

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>LIRO01</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q3468</u>
Instrument ID: <u>MSVOA_N</u>	Calibration Date(s): <u>10/23/2025</u> <u>10/23/2025</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>09:42</u> <u>12:02</u>
GC Column: <u>RXI-624</u> ID: <u>0.25</u> (mm)	

LAB FILE ID:		RRF001 = VN088078.D		RRF005 = VN088079.D		RRF020 = VN088080.D		
		RRF050 = VN088081.D		RRF100 = VN088082.D		RRF150 = VN088083.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Methyl tert-butyl Ether	2.356	2.498	2.493	1.790	2.370	2.342	2.308	11.4
Benzene	1.621	1.577	1.609	1.216	1.372	1.566	1.494	10.9
Toluene	0.971	0.958	0.991	0.755	0.862	0.989	0.921	10.2
Ethyl Benzene	2.069	2.066	2.123	1.649	1.873	2.177	1.993	9.9
m/p-Xylenes	0.733	0.774	0.804	0.635	0.708	0.827	0.747	9.4
o-Xylene	0.670	0.742	0.764	0.604	0.684	0.784	0.708	9.6
Isopropylbenzene	4.189	3.857	4.067	3.173	3.580	4.272	3.856	10.8
n-propylbenzene	4.874	4.855	4.930	3.925	4.436	5.217	4.706	9.7
1,3,5-Trimethylbenzene	3.329	3.254	3.385	2.669	3.028	3.551	3.203	9.8
tert-Butylbenzene	3.154	2.876	3.002	2.358	2.646	3.131	2.861	10.8
1,2,4-Trimethylbenzene	3.242	3.316	3.465	2.700	3.012	3.491	3.204	9.4
sec-Butylbenzene	4.578	4.273	4.382	3.490	3.881	4.616	4.203	10.4
p-Isopropyltoluene	3.269	3.541	3.602	2.866	3.170	3.783	3.372	9.9
n-Butylbenzene	3.396	3.438	3.493	2.827	3.158	3.730	3.340	9.3
Naphthalene	3.451	3.110	3.437	2.732	3.203	3.987	3.320	12.6
1,2-Dichloroethane-d4		0.914	0.915	0.799	0.822	0.872	0.865	6.1
Dibromofluoromethane		0.343	0.344	0.315	0.327	0.350	0.336	4.3
Toluene-d8		1.283	1.298	1.224	1.283	1.384	1.294	4.4
4-Bromofluorobenzene		0.435	0.457	0.439	0.466	0.488	0.457	4.8

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.