

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3468</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>LiRo Engineers, Inc.</u>	Datafile :	<u>VN088131.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1028WBS02	Methyl tert-butyl Ether	20	21.2	ug/L	106			78	114	
	Benzene	20	21.7	ug/L	109			82	109	
	Toluene	20	21.5	ug/L	108			82	110	
	Ethyl Benzene	20	22.0	ug/L	110		*	83	109	
	m/p-Xylenes	40	43.2	ug/L	108			82	110	
	o-Xylene	20	22.1	ug/L	111		*	83	109	
	Isopropylbenzene	20	21.3	ug/L	106			83	112	
	N-propylbenzene	20	21.7	ug/L	109			83	112	
	1,3,5-Trimethylbenzene	20	21.9	ug/L	110			85	112	
	tert-Butylbenzene	20	21.1	ug/L	106			83	112	
	1,2,4-Trimethylbenzene	20	22.0	ug/L	110			85	111	
	Sec-butylbenzene	20	21.6	ug/L	108			81	114	
	p-Isopropyltoluene	20	22.5	ug/L	113			78	116	
	n-Butylbenzene	20	21.5	ug/L	108			75	115	
	Naphthalene	20	20.8	ug/L	104			78	119	