

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3468</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>LiRo Engineers, Inc.</u>	Datafile :	<u>VN088150.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1029WBS01	Methyl tert-butyl Ether	20	19.6	ug/L	98			78	114	
	Benzene	20	19.2	ug/L	96			82	109	
	Toluene	20	18.6	ug/L	93			82	110	
	Ethyl Benzene	20	18.2	ug/L	91			83	109	
	m/p-Xylenes	40	35.7	ug/L	89			82	110	
	o-Xylene	20	18.3	ug/L	92			83	109	
	Isopropylbenzene	20	17.1	ug/L	86			83	112	
	N-propylbenzene	20	17.0	ug/L	85			83	112	
	1,3,5-Trimethylbenzene	20	17.8	ug/L	89			85	112	
	tert-Butylbenzene	20	16.2	ug/L	81		*	83	112	
	1,2,4-Trimethylbenzene	20	17.8	ug/L	89			85	111	
	Sec-butylbenzene	20	16.3	ug/L	81			81	114	
	p-Isopropyltoluene	20	17.2	ug/L	86			78	116	
	n-Butylbenzene	20	16.6	ug/L	83			75	115	
	Naphthalene	20	17.3	ug/L	86			78	119	