

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>LIRO01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3468</u>
Lab File ID:	<u>VN088127.D</u>	Date Analyzed:	<u>10/28/2025</u>
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>10:23</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	299407	8.21	575460	9.08	507990	11.85
UPPER LIMIT	598814	8.706	1150920	9.582	1015980	12.347
LOWER LIMIT	149704	7.706	287730	8.582	253995	11.347
EPA SAMPLE NO.						
MW-08	280779	8.21	636841	9.08	622260	11.85
MW-10	236497	8.21	529671	9.08	500860	11.85
MW-12	238860	8.21	531874	9.08	512440	11.85
VN1028WBL01	231261	8.21	516618	9.08	484454	11.85
VN1028WBS02	325166	8.21	590689	9.08	518318	11.85

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.