

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: LIRO01
 Lab Code: ACE SDG NO.: Q3468
 Lab File ID: VN088147.D Date Analyzed: 10/29/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:24
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	276791	8.21	513941	9.08	450607	11.85
UPPER LIMIT	553582	8.706	1027880	9.583	901214	12.347
LOWER LIMIT	138396	7.706	256971	8.583	225304	11.347
EPA SAMPLE NO.						
MW-06	239807	8.21	555035	9.08	540568	11.85
MW-08DL	214217	8.21	486300	9.08	458325	11.85
MW-13	328066	8.21	750626	9.08	708184	11.85
MW-12DL	208467	8.21	468936	9.08	447110	11.85
VN1029WBL01	220088	8.21	494622	9.08	463889	11.85
VN1029WBS01	234631	8.21	449642	9.08	401244	11.85

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.