

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: LIRO01
 Lab Code: ACE SDG NO.: Q3468
 Lab File ID: VN088102.D Date Analyzed: 10/27/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:40
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	267876	13.77				
UPPER LIMIT	535752	14.27				
LOWER LIMIT	133938	13.27				
EPA SAMPLE NO.						
MW-11	269618	13.77				
VN1027WBL01	199580	13.77				
VN1027WBS01	256562	13.77				
VN1027WBSD01	241641	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.