

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>LIRO01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3468</u>
Lab File ID:	<u>VN088147.D</u>	Date Analyzed:	<u>10/29/2025</u>
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>10:24</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	235546	13.77				
UPPER LIMIT	471092	14.27				
LOWER LIMIT	117773	13.27				
EPA SAMPLE NO.						
MW-06	270083	13.77				
MW-08DL	232108	13.77				
MW-13	354853	13.77				
MW-12DL	226296	13.77				
VN1029WBL01	222427	13.77				
VN1029WBS01	206462	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.