

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: LIRO01

Lab Code: ACE

SDG NO.: Q3468

Lab File ID: VN088101.D

BFB Injection Date: 10/27/2025

Instrument ID: MSVOA_N

BFB Injection Time: 09:58

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.7
75	30.0 - 60.0% of mass 95	57.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.1
173	Less than 2.0% of mass 174	0.4 (0.6) 1
174	50.0 - 100.0% of mass 95	68.6
175	5.0 - 9.0% of mass 174	5.1 (7.4) 1
176	95.0 - 101.0% of mass 174	65.3 (95.2) 1
177	5.0 - 9.0% of mass 176	4.2 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN088102.D	10/27/2025	10:40
VN1027WBL01	VN1027WBL01	VN088104.D	10/27/2025	11:34
VN1027WBS01	VN1027WBS01	VN088105.D	10/27/2025	11:55
VN1027WBSD01	VN1027WBSD01	VN088106.D	10/27/2025	12:29
MW-11	Q3468-04	VN088124.D	10/27/2025	18:41