

**TPH GC INITIAL CALIBRATION SUMMARY**

Lab Name: Alliance

Contract: SCIA01

ProjectID: 7 Rynda Road, Maplewood

Lab Code: ACE

SDG No.: Q3474

| Calibration Sequence : FF102925 |            | Test : TPH GC    |                 |
|---------------------------------|------------|------------------|-----------------|
| Concentration (PPM)             | Area Count | Reference Factor | File ID         |
| 85                              | 9472915    | 111446           | FF016603.D      |
| 170                             | 18165984   | 106859           | FF016604.D      |
| 340                             | 37023074   | 108891           | FF016605.D      |
| 850                             | 89698209   | 105527           | FF016606.D      |
| 1700                            | 169764046  | 99861            | FF016607.D      |
| AVG RF : 106517                 |            | % RSD : 4.074    | AVG RT : 15.059 |