

Surrogate Summary

SDG No.: Q3474

Client: Sciacca General Contractors, LLC

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3474-02	VOC	1,2-Dichloroethane-d4	50	58.2	116		63	155
		Dibromofluoromethane	50	52.6	105		70	134
		Toluene-d8	50	48.9	98		74	123
		4-Bromofluorobenzene	50	38.1	76		17	146
VY1028SBL01	VY1028SBL01	1,2-Dichloroethane-d4	50	50.7	101		63	155
		Dibromofluoromethane	50	50.0	100		70	134
		Toluene-d8	50	48.8	98		74	123
		4-Bromofluorobenzene	50	39.4	79		17	146
VY1028SBS01	VY1028SBS01	1,2-Dichloroethane-d4	50	50.3	101		63	155
		Dibromofluoromethane	50	52.3	105		70	134
		Toluene-d8	50	51.3	103		74	123
		4-Bromofluorobenzene	50	51.0	102		17	146
VY1028SBSD01	VY1028SBSD01	1,2-Dichloroethane-d4	50	49.3	99		63	155
		Dibromofluoromethane	50	51.3	103		70	134
		Toluene-d8	50	50.8	102		74	123
		4-Bromofluorobenzene	50	49.9	100		17	146