

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG NO.: Q3474
 Lab File ID: VY023598.D Date Analyzed: 10/28/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:02
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	477691	7.71	674925	8.62	592275	11.41
UPPER LIMIT	955382	8.213	1349850	9.115	1184550	11.914
LOWER LIMIT	238846	7.213	337463	8.115	296138	10.914
EPA SAMPLE NO.						
VOC	624082	7.71	1042505	8.62	846455	11.41
VY1028SBL01	751348	7.71	1250620	8.62	1019446	11.42
VY1028SBS01	466766	7.71	688703	8.62	608323	11.41
VY1028SBSD01	455764	7.71	666569	8.62	588558	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.