

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN102825

Review By	rahul	Review On	10/28/2025 3:41:07 PM
Supervise By	mohammad	Supervise On	10/30/2025 5:29:45 AM
SubDirectory	BN102825	HP Acquire Method	BNA_N, 8270_HP Processing Method BN102825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904
CCC	SP6902
Internal Standard/PEM	SP6830,1ul/100ul sample
ICV/I.BLK	SP6906
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038110.D	28 Oct 2025 12:05		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN038111.D	28 Oct 2025 13:00		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN038112.D	28 Oct 2025 13:36		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN038113.D	28 Oct 2025 14:13	compound #20 kept on LR	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN038114.D	28 Oct 2025 14:49		RC/JU	Ok,M
6	SSTDICC1.6	SSTDICC1.6	BN038115.D	28 Oct 2025 15:25		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN038116.D	28 Oct 2025 16:02		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN038117.D	28 Oct 2025 16:38	Compound #20 removed from 5PPM	RC/JU	Ok
9	SSTDICV0.4	ICVBN102825	BN038118.D	28 Oct 2025 17:51		RC/JU	Ok
10	PB170167BL	PB170167BL	BN038119.D	28 Oct 2025 19:04	END CCC missing, Analyzed for Contamination check.	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN103025

Review By	rahul	Review On	10/31/2025 2:18:33 PM
Supervise By	Jagrut	Supervise On	10/31/2025 2:25:45 PM
SubDirectory	BN103025	HP Acquire Method	BNA_N, 8270_HP Processing Method BN102825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904
CCC	SP6902
Internal Standard/PEM	SP6830,1ul/100ul sample
ICV/I.BLK	SP6906
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038120.D	30 Oct 2025 09:21		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN038121.D	30 Oct 2025 10:01	Comp#14 Surr,24 failed from lowside	RC/JU	Ok
3	PB170298BL	PB170298BL	BN038122.D	30 Oct 2025 10:49		RC/JU	Ok
4	PB170298BS	PB170298BS	BN038123.D	30 Oct 2025 11:25		RC/JU	Ok
5	Q3478-01	38072-061125	BN038124.D	30 Oct 2025 13:12	PT Sample, Need 50X Dilution	RC/JU	Dilution
6	Q3478-01DL	38072-061125DL	BN038125.D	30 Oct 2025 16:08		RC/JU	Ok
7	Q3015-21	WP0925-RR-PAH-WP	BN038126.D	30 Oct 2025 17:12	Already analyzed with OK Status	RC/JU	Confirms
8	Q3015-21DL	WP0925-RR-PAH-WPDL	BN038127.D	30 Oct 2025 17:48	Already analyzed with OK Status	RC/JU	Confirms
9	SSTDCCC0.4	SSTDCCC0.4EC	BN038128.D	30 Oct 2025 18:24		RC/JU	Ok

M : Manual Integration