

Surrogate Summary

SW-846

SDG No.: Q3478

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB170298BL	PB170298BL	2-Methylnaphthalene-d10	0.4	0.33	83		20	139
		Fluoranthene-d10	0.4	0.33	83		54	157
		Nitrobenzene-d5	0.4	0.36	89		27	154
		2-Fluorobiphenyl	0.4	0.41	101		30	155
		Terphenyl-d14	0.4	0.45	113		54	175
PB170298BS	PB170298BS	2-Methylnaphthalene-d10	0.4	0.36	91		20	139
		Fluoranthene-d10	0.4	0.33	81		54	157
		Nitrobenzene-d5	0.4	0.37	91		27	154
		2-Fluorobiphenyl	0.4	0.41	102		30	155
		Terphenyl-d14	0.4	0.44	109		54	175
Q3478-01	38072-061125	2-Methylnaphthalene-d10	0.4	0.38	94		20	139
		Fluoranthene-d10	0.4	0.37	93		54	157
		Nitrobenzene-d5	0.4	0.38	95		27	154
		2-Fluorobiphenyl	0.4	0.42	106		30	155
		Terphenyl-d14	0.4	0.41	102		54	175
Q3478-01DL	38072-061125DL	2-Methylnaphthalene-d10	0.4	0.35	88		20	139
		Fluoranthene-d10	0.4	0.35	88		54	157
		Nitrobenzene-d5	0.4	0.60	150		27	154
		2-Fluorobiphenyl	0.4	0.65	162	*	30	155
		Terphenyl-d14	0.4	0.45	113		54	175