

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3478

Client ID : SSTDCCC0.4

Date Analyzed: 10/30/2025

Lab File ID: BN038121.D

Time Analyzed: 10:01

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	9463	7.775	27993	10.58	15485	14.44
UPPER LIMIT	18926	8.275	55986	11.084	30970	14.941
LOWER LIMIT	4731.5	7.275	13996.5	10.084	7742.5	13.941
EPA SAMPLE NO.						
01 PB170298BL	9151	7.79	25373	10.59	13033	14.44
02 PB170298BS	11299	7.78	32410	10.58	16943	14.44
03 38072-061125	8933	7.78	29561	10.58	16222	14.44
04 38072-061125DL	8055	7.79	24698	10.59	13761	14.44

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.