



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Alliance Technical Group, LLC - Newark

Project Name: NJ Soil PT

Project # N/A

Order ID # Q3483

Test Name: Ammonia,Anions Group1,Anions Group2,Corrosivity,Cyanide,Flash Point,Hexavalent Chromium,Oil and Grease,Phosphorus, Total,TKN,TOC,TS

A. Number of Samples and Date of Receipt:

11 Solid samples were received on 10/29/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Ammonia,Anions Group1,Anions Group2,Corrosivity,Cyanide,Flash Point,Hexavalent Chromium,Oil and Grease,Phosphorus, Total,TKN,TOC,TS. This data package contains results for Ammonia,Anions Group1,Anions Group2,Corrosivity,Cyanide,Flash Point,Hexavalent Chromium,Oil and Grease,Phosphorus, Total,TKN,TOC,TS.

C. Analytical Techniques:

The analysis of Flash Point was based on method 1010B, The analysis of Hexavalent Chromium was based on method 7196A, The analysis of Cyanide was based on method 9012B, The analysis of Cyanide was based on method 9014, The analysis of Corrosivity was based on method 9045D, The analysis of Anions Group1,Anions Group2 was based on method 9056A, The analysis of TOC was based on method 9060A, The analysis of Oil and Grease was based on method 9071B, The analysis of TOC was based on method Lloyd Kahn, The analysis of TS was based on method SM2540 B, The analysis of TKN was based on method SM4500 N Org B or C and The analysis of Ammonia was based on method SM4500-NH3.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

Sample HW1025-PT-AN-SOIL was diluted due to high concentrations for Chloride, Nitrate, Sulfate & Sample HW1025-PT-CN-SOIL was diluted due to high concentrations for Cyanide & Sample HW1025-PT-CN-SOIL was diluted due to high concentrations for Cyanide & Sample HW1025-PT-CR6-SOIL was diluted due to high concentrations for Hexavalent Chromium & Sample HW1025-PT-NUT-SOIL was diluted due to high concentrations for Ammonia as N, TKN, Phosphorus, Total & Sample HW1025-PT-NO2-SOIL was diluted due to high concentrations for Nitrite.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Matrix Spike (SED-2MS) analysis met criteria for all compounds except for TKN due to Sample matrix interference.



The Matrix Spike Duplicate (SED-2MSD) analysis met criteria for all compounds except for TKN due to Sample matrix interference.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

Lab has used least representable sample weight for the sample Q3483-07 and Q3483-08 for TOC analysis. Therefore Lab has reported the TOC result with "OR" qualifier.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____