

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB138042

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	11/25/2025	11:15
2	Q3483-05		1	8	43.00	109.4	109.4	11/25/2025	11:45
3	Q3708-01		1	12	100.00	>212.0	>212.0	11/25/2025	12:15
4	Q3708-01DUP		1	12	100.00	>212.0	>212.0	11/25/2025	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6138042

WorkList Name : FP-11-24.

WorkList ID : 193329

Department : Wet-Chemistry

Date : 11-24-2025 16:38:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3483-05	HW1025-PT-FP-SOIL	Solid	Flash Point	Cool 4 deg C	ALLI03	QA Of	10/27/2025	1010B
Q3708-01	112025	Solid	Flash Point	Cool 4 deg C	PSEG03	D41	11/21/2025	1010B

Date/Time 11/25/2025 11:00
 Raw Sample Received by: RM LWC
 Raw Sample Relinquished by: JG LWC

Date/Time 11/25/2025 13:25
 Raw Sample Received by: JG LWC
 Raw Sample Relinquished by: RM LWC