

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q3483

Contract: ALLI03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170330BS							
Aluminum	mg/Kg	1000	989		99	80 - 120	P
Antimony	mg/Kg	50.0	52.4		105	80 - 120	P
Arsenic	mg/Kg	50.0	51.2		102	80 - 120	P
Barium	mg/Kg	250	253		101	80 - 120	P
Beryllium	mg/Kg	50.0	50.7		101	80 - 120	P
Cadmium	mg/Kg	50.0	50.8		102	80 - 120	P
Calcium	mg/Kg	5000	5130		103	80 - 120	P
Chromium	mg/Kg	50.0	51.0		102	80 - 120	P
Cobalt	mg/Kg	50.0	51.8		104	80 - 120	P
Copper	mg/Kg	500	516		103	80 - 120	P
Iron	mg/Kg	2500	2690		108	80 - 120	P
Lead	mg/Kg	250	248		99	80 - 120	P
Magnesium	mg/Kg	5000	5000		100	80 - 120	P
Manganese	mg/Kg	500	506		101	80 - 120	P
Nickel	mg/Kg	50.0	51.8		104	80 - 120	P
Potassium	mg/Kg	2500	2550		102	80 - 120	P
Tin	mg/Kg	50.0	51.2		102	80 - 120	P
Molybdenum	mg/Kg	500	503		101	80 - 120	P
Selenium	mg/Kg	50.0	51.6		103	80 - 120	P
Silver	mg/Kg	50.0	51.5		103	80 - 120	P
Sodium	mg/Kg	5000	5090		102	80 - 120	P
Thallium	mg/Kg	50.0	50.8		102	80 - 120	P
Vanadium	mg/Kg	50.0	50.6		101	80 - 120	P
Zinc	mg/Kg	500	527		105	80 - 120	P
Strontium	mg/Kg	50.0	50.6		101	80 - 120	P