

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB102725

Review By	yogesh	Review On	10/29/2025 7:48:30 AM		
Supervise By	mohammad	Supervise On	10/29/2025 7:50:23 AM		
SubDirectory	FB102725	HP Acquire Method	FID_B	HP Processing Method	fb102725.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24717,PP25006,PP25007,PP25008,PP25009,PP25010
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24718,PP25011

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	5 GRO STD		FB032267.D	27 Oct 2025 9:43		YP/AJ	Ok
2	10 GRO STD		FB032268.D	27 Oct 2025 10:11		YP/AJ	Ok
3	20 GRO STD		FB032269.D	27 Oct 2025 10:38		YP/AJ	Ok
4	50 GRO STD		FB032270.D	27 Oct 2025 11:06		YP/AJ	Ok
5	100 GRO STD		FB032271.D	27 Oct 2025 11:34		YP/AJ	Ok
6	FB1027GROICV		FB032272.D	27 Oct 2025 12:20		YP/AJ	Ok

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB110425

Review By	Amit	Review On	11/5/2025 12:48:53 PM
Supervise By	yogesh	Supervise On	11/5/2025 2:53:07 PM
SubDirectory	FB110425	HP Acquire Method	FID_B
		HP Processing Method	fb102725

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24717,PP25006,PP25007,PP25008,PP25009,PP25010
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP25017,PP25018

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB032285.D	4 Nov 2025 8:23		YP/AJ	Ok
2	VBF1104S1		FB032286.D	4 Nov 2025 9:14		YP/AJ	Ok
3	VBF1104S2		FB032287.D	4 Nov 2025 9:42		YP/AJ	Ok
4	BSF1104S1		FB032288.D	4 Nov 2025 10:10		YP/AJ	Ok
5	BSF1104S2		FB032289.D	4 Nov 2025 10:37		YP/AJ	Ok
6	Q3483-15		FB032290.D	4 Nov 2025 11:23	need 1000x	YP/AJ	Dilution
7	I.BLK		FB032291.D	4 Nov 2025 11:51		YP/AJ	Ok
8	Q3483-15		FB032292.D	4 Nov 2025 12:18	need further dilution	YP/AJ	Dilution
9	Q3483-15		FB032293.D	4 Nov 2025 12:46	surrogate fail	YP/AJ	Ok
10	Q3483-15		FB032294.D	4 Nov 2025 13:13	not required	YP/AJ	Not Ok
11	I.BLK		FB032295.D	4 Nov 2025 13:41		YP/AJ	Ok
12	20 PPB GRO STD		FB032296.D	4 Nov 2025 14:09		YP/AJ	Ok
13	Q3530-01		FB032297.D	4 Nov 2025 15:27	1 PPB SOIL LOD	YP/AJ	Not Ok
14	Q3530-01		FB032298.D	4 Nov 2025 15:55	4 PPB SOIL LOD	YP/AJ	Ok
15	Q3530-02		FB032299.D	4 Nov 2025 16:22	5 PPB SOIL LOQ	YP/AJ	Ok
16	20 PPB GRO STD		FB032300.D	4 Nov 2025 16:50		YP/AJ	Ok

M : Manual Integration