



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

HW1025-PT-PCBO-SOIL

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Lab Sample ID: Q3483-19

Date(s) Analyzed: 12/05/2025 12/05/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115623.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1242	1	5.482	5.432	5.532	6260	7100	15.15
COLUMN 1	2	5.504	5.454	5.554	7000		
	3	5.566	5.516	5.616	7830		
	4	5.661	5.611	5.711	7360		
	5	6.387	6.337	6.437	6870		
	1	4.724	4.674	4.774	6410		
COLUMN 2	2	4.743	4.693	4.793	6860	6100	
	3	4.916	4.866	4.966	5980		
	4	5.001	4.951	5.051	5920		
	5	5.519	5.469	5.569	5520		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

HW1025-PT-PCB-SOIL

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Lab Sample ID: Q3483-18

Date(s) Analyzed: 12/08/2025 12/08/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115637.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1254	1	6.316	6.266	6.366	7220	7900	19.44
COLUMN 1	2	6.532	6.482	6.582	7530		
	3	6.892	6.842	6.942	7850		
	4	7.174	7.124	7.224	8270		
	5	7.589	7.539	7.639	8560		
	1	5.509	5.459	5.559	5940		
COLUMN 2	2	5.657	5.607	5.707	6090	6500	
	3	6.058	6.008	6.108	6520		
	4	6.286	6.236	6.336	6700		
	5	6.702	6.652	6.752	7150		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

HW1025-PT-PCB-SOILD

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Lab Sample ID: Q3483-18DL

Date(s) Analyzed: 12/08/2025 12/08/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115639.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1254	1	6.317	6.267	6.367	7270	7500	5.48
COLUMN 1	2	6.533	6.483	6.583	7410		
	3	6.892	6.842	6.942	7370		
	4	7.175	7.125	7.225	7630		
	5	7.59	7.54	7.64	7610		
	1	5.51	5.46	5.56	6860		
COLUMN 2	2	5.658	5.608	5.708	6980	7100	
	3	6.058	6.008	6.108	7040		
	4	6.286	6.236	6.336	7140		
	5	6.703	6.653	6.753	7300		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB170817BS

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Lab Sample ID: PB170817BS

Date(s) Analyzed: 12/04/2025 12/04/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP076810.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.768	5.718	5.818	3810	3800	0
COLUMN 1	2	5.789	5.739	5.839	3890		
	3	5.852	5.802	5.902	3830		
	4	5.949	5.899	5.999	3820		
	5	6.241	6.191	6.291	3710		
	1	4.849	4.799	4.899	3810		
COLUMN 2	2	4.866	4.816	4.916	3880	3800	
	3	5.042	4.992	5.092	3860		
	4	5.083	5.033	5.133	3770		
	5	5.296	5.246	5.346	3700		
Aroclor-1260	1	7.358	7.308	7.408	3900	3600	2.74
COLUMN 1	2	7.611	7.561	7.661	3810		
	3	7.969	7.919	8.019	3290		
	4	8.193	8.143	8.243	3470		
	5	8.519	8.469	8.569	3480		
	1	6.323	6.273	6.373	3990		
COLUMN 2	2	6.51	6.46	6.56	3960	3700	
	3	6.662	6.612	6.712	3910		
	4	7.131	7.081	7.181	3420		
	5	7.372	7.322	7.422	3420		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB170627BS

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Lab Sample ID: PB170627BS

Date(s) Analyzed: 11/19/2025 11/19/2025

Instrument ID (1): ECD_Q

Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071335.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.497	4.447	4.547	130	130	9.52
COLUMN 1	2	4.516	4.466	4.566	132		
	3	4.573	4.523	4.623	129		
	4	4.664	4.614	4.714	131		
	5	4.949	4.899	4.999	128		
	1	3.87	3.82	3.92	145		
COLUMN 2	2	3.886	3.836	3.936	148	143	
	3	4.048	3.998	4.098	145		
	4	4.097	4.047	4.147	140		
	5	4.294	4.244	4.344	140		
Aroclor-1260	1	6.043	5.993	6.093	129	123	13.64
COLUMN 1	2	6.3	6.25	6.35	135		
	3	6.653	6.603	6.703	113		
	4	6.867	6.817	6.917	119		
	5	7.174	7.124	7.224	118		
	1	5.295	5.245	5.345	147	141	
COLUMN 2	2	5.484	5.434	5.534	149		
	3	5.627	5.577	5.677	149		
	4	6.091	6.041	6.141	129		
	5	6.336	6.286	6.386	129		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

EO-01-11182025MS

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Lab Sample ID: Q3669-01MS

Date(s) Analyzed: 11/19/2025 11/19/2025

Instrument ID (1): ECD_Q

Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071340.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.496	4.446	4.546	91.8	90.2	11.3
COLUMN 1	2	4.515	4.465	4.565	83.7		
	3	4.572	4.522	4.622	91.5		
	4	4.662	4.612	4.712	97.5		
	5	4.948	4.898	4.998	86.6		
	1	3.869	3.819	3.919	96.5		
COLUMN 2	2	3.886	3.836	3.936	97.2	101	
	3	4.048	3.998	4.098	106		
	4	4.096	4.046	4.146	100		
	5	4.294	4.244	4.344	107		
Aroclor-1260	1	6.042	5.992	6.092	87.8	86.2	1.17
COLUMN 1	2	6.299	6.249	6.349	110		
	3	6.652	6.602	6.702	76.5		
	4	6.867	6.817	6.917	77.8		
	5	7.174	7.124	7.224	78.5		
	1	5.296	5.246	5.346	93.4		
COLUMN 2	2	5.485	5.435	5.535	98.8	85.2	
	3	5.627	5.577	5.677	90.0		
	4	6.092	6.042	6.142	71.4		
	5	6.337	6.287	6.387	72.4		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

EO-01-11182025MSD

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Lab Sample ID: Q3669-01MSD

Date(s) Analyzed: 11/19/2025 11/19/2025

Instrument ID (1): ECD_Q

Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071341.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.496	4.446	4.546	93.0	91.7	9.65
COLUMN 1	2	4.515	4.465	4.565	88.2		
	3	4.573	4.523	4.623	92.2		
	4	4.663	4.613	4.713	96.6		
	5	4.948	4.898	4.998	88.6		
	1	3.87	3.82	3.92	90.2		
COLUMN 2	2	3.886	3.836	3.936	98.4	101	
	3	4.048	3.998	4.098	107		
	4	4.096	4.046	4.146	105		
	5	4.294	4.244	4.344	105		
Aroclor-1260	1	6.042	5.992	6.092	86.3	85.4	0.47
COLUMN 1	2	6.299	6.249	6.349	110		
	3	6.652	6.602	6.702	76.9		
	4	6.867	6.817	6.917	76.0		
	5	7.174	7.124	7.224	77.2		
	1	5.295	5.245	5.345	94.1		
COLUMN 2	2	5.484	5.434	5.534	99.0	85.0	
	3	5.627	5.577	5.677	89.0		
	4	6.092	6.042	6.142	73.9		
	5	6.336	6.286	6.386	69.2		