

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: ALLI03

Lab Code: ACE SDG NO.: Q3483

Instrument ID: ECD_P Date(s) Analyzed: 11/21/2025 11/21/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.82	4.72	4.92	25833800
		2	4.90	4.80	5.00	19927200
		3	4.98	4.88	5.08	55758000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.98	4.88	5.08	43472600
		2	5.51	5.41	5.61	22364600
		3	5.80	5.70	5.90	43924600
		4	5.96	5.86	6.06	22708400
		5	6.04	5.94	6.14	16825100
Aroclor-1262	500	1	8.20	8.10	8.30	98956600
		2	8.52	8.42	8.62	172451000
		3	8.84	8.74	8.94	125724000
		4	8.93	8.83	9.03	99218000
		5	9.59	9.49	9.69	71040600

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance
Contract: ALLI03

Lab Code: ACE
SDG NO.: Q3483

Instrument ID: ECD_P
Date(s) Analyzed: 11/21/2025 11/21/2025

GC Column: ZB-MR2
ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.98	3.88	4.08	26717800
		2	4.07	3.97	4.17	19991600
		3	4.15	4.05	4.25	60271800
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.15	4.05	4.25	47081000
		2	4.87	4.77	4.97	46115400
		3	5.05	4.95	5.15	24241400
		4	5.13	5.03	5.23	21532600
		5	5.30	5.20	5.40	23536200
Aroclor-1262	500	1	6.88	6.78	6.98	133315000
		2	7.14	7.04	7.24	119918000
		3	7.66	7.56	7.76	98515200
		4	7.72	7.62	7.82	174485000
		5	8.22	8.12	8.32	80357800