

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Continuing Calib Date: 12/05/2025

Initial Calibration Date(s): 12/04/2025

12/04/2025

Continuing Calib Time: 10:40

Initial Calibration Time(s): 10:43

21:19

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.47	5.48	5.38	5.58	0.01
Aroclor-1242-2 (2)	5.49	5.50	5.40	5.60	0.01
Aroclor-1242-3 (3)	5.56	5.56	5.46	5.66	0.00
Aroclor-1242-4 (4)	5.65	5.65	5.55	5.75	0.00
Aroclor-1242-5 (5)	6.38	6.38	6.28	6.48	0.00
Tetrachloro-m-xylene	4.33	4.34	4.24	4.44	0.01
Decachlorobiphenyl	9.91	9.91	9.81	10.01	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Continuing Calib Date: 12/05/2025

Initial Calibration Date(s): 12/04/2025

12/04/2025

Continuing Calib Time: 10:40

Initial Calibration Time(s): 10:43

21:19

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.71	4.72	4.62	4.82	0.01
Aroclor-1242-2 (2)	4.73	4.73	4.63	4.83	0.00
Aroclor-1242-3 (3)	4.91	4.91	4.81	5.01	0.00
Aroclor-1242-4 (4)	4.99	4.99	4.89	5.09	0.00
Aroclor-1242-5 (5)	5.51	5.51	5.41	5.61	0.00
Tetrachloro-m-xylene	3.64	3.64	3.54	3.74	0.00
Decachlorobiphenyl	8.62	8.62	8.52	8.72	0.00