

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Continuing Calib Date: 12/05/2025

Initial Calibration Date(s): 12/04/2025

12/04/2025

Continuing Calib Time: 11:16

Initial Calibration Time(s): 10:43

21:19

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.32	6.32	6.22	6.42	0.00
Aroclor-1254-2	(2)	6.53	6.53	6.43	6.63	0.00
Aroclor-1254-3	(3)	6.89	6.89	6.79	6.99	0.00
Aroclor-1254-4	(4)	7.17	7.17	7.07	7.27	0.00
Aroclor-1254-5	(5)	7.59	7.59	7.49	7.69	0.00
Tetrachloro-m-xylene		4.33	4.33	4.23	4.43	0.00
Decachlorobiphenyl		9.91	9.91	9.81	10.01	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Continuing Calib Date: 12/05/2025

Initial Calibration Date(s): 12/04/2025

12/04/2025

Continuing Calib Time: 11:16

Initial Calibration Time(s): 10:43

21:19

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.51	5.51	5.41	5.61	0.00
Aroclor-1254-2 (2)	5.66	5.66	5.56	5.76	0.00
Aroclor-1254-3 (3)	6.06	6.06	5.96	6.16	0.00
Aroclor-1254-4 (4)	6.29	6.29	6.19	6.39	0.00
Aroclor-1254-5 (5)	6.70	6.70	6.60	6.80	0.00
Tetrachloro-m-xylene	3.64	3.64	3.54	3.74	0.00
Decachlorobiphenyl	8.62	8.62	8.52	8.72	0.00