

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Continuing Calib Date: 12/05/2025

Initial Calibration Date(s): 12/04/2025

12/04/2025

Continuing Calib Time: 11:35

Initial Calibration Time(s): 10:43

21:19

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1268-1 (1)	8.49	8.50	8.40	8.60	0.01
Aroclor-1268-2 (2)	8.58	8.58	8.48	8.68	0.00
Aroclor-1268-3 (3)	8.80	8.80	8.70	8.90	0.00
Aroclor-1268-4 (4)	9.20	9.20	9.10	9.30	0.00
Aroclor-1268-5 (5)	9.59	9.59	9.49	9.69	0.00
Tetrachloro-m-xylene	4.34	4.34	4.24	4.44	0.00
Decachlorobiphenyl	9.91	9.91	9.81	10.01	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Continuing Calib Date: 12/05/2025

Initial Calibration Date(s): 12/04/2025

12/04/2025

Continuing Calib Time: 11:35

Initial Calibration Time(s): 10:43

21:19

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1268-1 (1)	7.52	7.52	7.42	7.62	0.00
Aroclor-1268-2 (2)	7.59	7.59	7.49	7.69	0.00
Aroclor-1268-3 (3)	7.79	7.80	7.70	7.90	0.01
Aroclor-1268-4 (4)	8.08	8.08	7.98	8.18	0.00
Aroclor-1268-5 (5)	8.37	8.37	8.27	8.47	0.00
Tetrachloro-m-xylene	3.64	3.64	3.54	3.74	0.00
Decachlorobiphenyl	8.62	8.62	8.52	8.72	0.00