

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Continuing Calib Date: 12/08/2025

Initial Calibration Date(s): 12/04/2025

12/04/2025

Continuing Calib Time: 09:58

Initial Calibration Time(s): 10:43

21:19

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.48	5.48	5.38	5.58	0.00
Aroclor-1016-2	(2)	5.50	5.50	5.40	5.60	0.00
Aroclor-1016-3	(3)	5.56	5.56	5.46	5.66	0.00
Aroclor-1016-4	(4)	5.66	5.65	5.55	5.75	-0.01
Aroclor-1016-5	(5)	5.95	5.95	5.85	6.05	0.00
Aroclor-1260-1	(1)	7.06	7.06	6.96	7.16	0.00
Aroclor-1260-2	(2)	7.32	7.31	7.21	7.41	-0.01
Aroclor-1260-3	(3)	7.67	7.67	7.57	7.77	0.00
Aroclor-1260-4	(4)	7.90	7.89	7.79	7.99	-0.01
Aroclor-1260-5	(5)	8.20	8.20	8.10	8.30	0.00
Tetrachloro-m-xylene		4.34	4.34	4.24	4.44	0.00
Decachlorobiphenyl		9.92	9.91	9.81	10.01	-0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Continuing Calib Date: 12/08/2025

Initial Calibration Date(s): 12/04/2025

12/04/2025

Continuing Calib Time: 09:58

Initial Calibration Time(s): 10:43

21:19

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.71	4.72	4.62	4.82	0.01
Aroclor-1016-2	(2)	4.73	4.73	4.63	4.83	0.00
Aroclor-1016-3	(3)	4.91	4.91	4.81	5.01	0.00
Aroclor-1016-4	(4)	4.95	4.95	4.85	5.05	0.00
Aroclor-1016-5	(5)	5.16	5.16	5.06	5.26	0.00
Aroclor-1260-1	(1)	6.19	6.19	6.09	6.29	0.00
Aroclor-1260-2	(2)	6.38	6.38	6.28	6.48	0.00
Aroclor-1260-3	(3)	6.53	6.53	6.43	6.63	0.00
Aroclor-1260-4	(4)	7.00	7.00	6.90	7.10	0.00
Aroclor-1260-5	(5)	7.24	7.24	7.14	7.34	0.00
Tetrachloro-m-xylene		3.64	3.64	3.54	3.74	0.00
Decachlorobiphenyl		8.62	8.62	8.52	8.72	0.00