

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Continuing Calib Date: 12/08/2025

Initial Calibration Date(s): 12/04/2025

12/04/2025

Continuing Calib Time: 20:22

Initial Calibration Time(s): 10:43

21:19

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.47	5.47	5.37	5.57	0.00
Aroclor-1248-2 (2)	5.74	5.75	5.65	5.85	0.01
Aroclor-1248-3 (3)	5.94	5.95	5.85	6.05	0.01
Aroclor-1248-4 (4)	6.34	6.34	6.24	6.44	0.00
Aroclor-1248-5 (5)	6.38	6.38	6.28	6.48	0.00
Tetrachloro-m-xylene	4.33	4.34	4.24	4.44	0.01
Decachlorobiphenyl	9.91	9.91	9.81	10.01	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Continuing Calib Date: 12/08/2025

Initial Calibration Date(s): 12/04/2025

12/04/2025

Continuing Calib Time: 20:22

Initial Calibration Time(s): 10:43

21:19

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.71	4.71	4.61	4.81	0.00
Aroclor-1248-2 (2)	4.95	4.95	4.85	5.05	0.00
Aroclor-1248-3 (3)	4.99	4.99	4.89	5.09	0.00
Aroclor-1248-4 (4)	5.16	5.16	5.06	5.26	0.00
Aroclor-1248-5 (5)	5.55	5.55	5.45	5.65	0.00
Tetrachloro-m-xylene	3.64	3.64	3.54	3.74	0.00
Decachlorobiphenyl	8.62	8.62	8.52	8.72	0.00