

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Continuing Calib Date: 12/04/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 09:43

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.77	5.78	5.68	5.88	0.01
Aroclor-1248-2 (2)	6.04	6.05	5.95	6.15	0.01
Aroclor-1248-3 (3)	6.24	6.25	6.15	6.35	0.01
Aroclor-1248-4 (4)	6.64	6.65	6.55	6.75	0.01
Aroclor-1248-5 (5)	6.68	6.69	6.59	6.79	0.01
Tetrachloro-m-xylene	4.62	4.63	4.53	4.73	0.01
Decachlorobiphenyl	10.37	10.38	10.28	10.48	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Continuing Calib Date: 12/04/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 09:43

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.85	4.86	4.76	4.96	0.01
Aroclor-1248-2 (2)	5.08	5.09	4.99	5.19	0.01
Aroclor-1248-3 (3)	5.13	5.13	5.03	5.23	0.00
Aroclor-1248-4 (4)	5.30	5.30	5.20	5.40	0.00
Aroclor-1248-5 (5)	5.69	5.69	5.59	5.79	0.00
Tetrachloro-m-xylene	3.77	3.78	3.68	3.88	0.01
Decachlorobiphenyl	8.77	8.77	8.67	8.87	0.00