

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/04/2025 12/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115617.D **Time Analyzed:** 10:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.473	5.375	5.575	556.180	500.000	11.2
Aroclor-1016-2	5.495	5.396	5.596	561.020	500.000	12.2
Aroclor-1016-3	5.557	5.458	5.658	568.330	500.000	13.7
Aroclor-1016-4	5.652	5.553	5.753	567.520	500.000	13.5
Aroclor-1016-5	5.944	5.846	6.046	580.580	500.000	16.1
Aroclor-1260-1	7.056	6.958	7.158	567.230	500.000	13.4
Aroclor-1260-2	7.310	7.212	7.412	542.340	500.000	8.5
Aroclor-1260-3	7.668	7.570	7.770	604.840	500.000	21.0
Aroclor-1260-4	7.889	7.791	7.991	558.600	500.000	11.7
Aroclor-1260-5	8.195	8.098	8.298	554.310	500.000	10.9
Decachlorobiphenyl	9.909	9.812	10.012	56.410	50.000	12.8
Tetrachloro-m-xylene	4.335	4.236	4.436	55.780	50.000	11.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/04/2025 12/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115617.D **Time Analyzed:** 10:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.713	4.615	4.815	522.110	500.000	4.4
Aroclor-1016-2	4.731	4.634	4.834	525.230	500.000	5.0
Aroclor-1016-3	4.905	4.808	5.008	520.570	500.000	4.1
Aroclor-1016-4	4.947	4.851	5.051	496.070	500.000	-0.8
Aroclor-1016-5	5.159	5.062	5.262	578.100	500.000	15.6
Aroclor-1260-1	6.188	6.091	6.291	538.380	500.000	7.7
Aroclor-1260-2	6.375	6.278	6.478	528.290	500.000	5.7
Aroclor-1260-3	6.528	6.431	6.631	497.790	500.000	-0.4
Aroclor-1260-4	6.998	6.901	7.101	528.370	500.000	5.7
Aroclor-1260-5	7.240	7.142	7.342	527.290	500.000	5.5
Decachlorobiphenyl	8.619	8.522	8.722	52.580	50.000	5.2
Tetrachloro-m-xylene	3.637	3.538	3.738	52.230	50.000	4.5