

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/04/2025 12/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 12/05/2025
Lab Sample No.: AR1242CCC500 **Data File :** PO115618.D **Time Analyzed:** 10:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1242-1	5.473	5.375	5.575	545.650	500.000	9.1
Aroclor-1242-2	5.494	5.396	5.596	565.190	500.000	13.0
Aroclor-1242-3	5.556	5.458	5.658	555.760	500.000	11.2
Aroclor-1242-4	5.652	5.554	5.754	560.300	500.000	12.1
Aroclor-1242-5	6.377	6.280	6.480	548.780	500.000	9.8
Decachlorobiphenyl	9.908	9.811	10.011	53.250	50.000	6.5
Tetrachloro-m-xylene	4.334	4.236	4.436	52.990	50.000	6.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/04/2025 12/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 12/05/2025
Lab Sample No.: AR1242CCC500 **Data File :** PO115618.D **Time Analyzed:** 10:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1242-1	4.712	4.615	4.815	535.250	500.000	7.1
Aroclor-1242-2	4.731	4.634	4.834	518.640	500.000	3.7
Aroclor-1242-3	4.905	4.808	5.008	522.570	500.000	4.5
Aroclor-1242-4	4.989	4.892	5.092	523.220	500.000	4.6
Aroclor-1242-5	5.509	5.412	5.612	553.200	500.000	10.6
Decachlorobiphenyl	8.619	8.522	8.722	49.460	50.000	-1.1
Tetrachloro-m-xylene	3.637	3.538	3.738	51.910	50.000	3.8