

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/04/2025 12/04/2025

Client Sample No.: CCAL03 **Date Analyzed:** 12/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115624.D **Time Analyzed:** 14:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.473	5.375	5.575	577.390	500.000	15.5
Aroclor-1016-2	5.495	5.396	5.596	575.540	500.000	15.1
Aroclor-1016-3	5.557	5.458	5.658	585.310	500.000	17.1
Aroclor-1016-4	5.653	5.553	5.753	584.550	500.000	16.9
Aroclor-1016-5	5.944	5.846	6.046	578.950	500.000	15.8
Aroclor-1260-1	7.058	6.958	7.158	578.670	500.000	15.7
Aroclor-1260-2	7.311	7.212	7.412	562.560	500.000	12.5
Aroclor-1260-3	7.668	7.570	7.770	563.920	500.000	12.8
Aroclor-1260-4	7.890	7.791	7.991	578.380	500.000	15.7
Aroclor-1260-5	8.197	8.098	8.298	572.300	500.000	14.5
Decachlorobiphenyl	9.912	9.812	10.012	57.360	50.000	14.7
Tetrachloro-m-xylene	4.334	4.236	4.436	57.230	50.000	14.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/04/2025 12/04/2025

Client Sample No.: CCAL03 **Date Analyzed:** 12/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115624.D **Time Analyzed:** 14:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.712	4.615	4.815	528.170	500.000	5.6
Aroclor-1016-2	4.731	4.634	4.834	537.160	500.000	7.4
Aroclor-1016-3	4.905	4.808	5.008	529.790	500.000	6.0
Aroclor-1016-4	4.947	4.851	5.051	488.880	500.000	-2.2
Aroclor-1016-5	5.159	5.062	5.262	595.630	500.000	19.1
Aroclor-1260-1	6.188	6.091	6.291	557.340	500.000	11.5
Aroclor-1260-2	6.376	6.278	6.478	541.030	500.000	8.2
Aroclor-1260-3	6.528	6.431	6.631	507.850	500.000	1.6
Aroclor-1260-4	6.999	6.901	7.101	535.850	500.000	7.2
Aroclor-1260-5	7.241	7.142	7.342	530.850	500.000	6.2
Decachlorobiphenyl	8.620	8.522	8.722	53.320	50.000	6.6
Tetrachloro-m-xylene	3.636	3.538	3.738	52.860	50.000	5.7