

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/04/2025 12/04/2025

Client Sample No.: CCAL04 **Date Analyzed:** 12/05/2025
Lab Sample No.: AR1242CCC500 **Data File :** PO115625.D **Time Analyzed:** 15:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1242-1	5.473	5.375	5.575	563.830	500.000	12.8
Aroclor-1242-2	5.494	5.396	5.596	579.980	500.000	16.0
Aroclor-1242-3	5.555	5.458	5.658	572.740	500.000	14.5
Aroclor-1242-4	5.651	5.554	5.754	562.690	500.000	12.5
Aroclor-1242-5	6.377	6.280	6.480	583.060	500.000	16.6
Decachlorobiphenyl	9.909	9.811	10.011	54.190	50.000	8.4
Tetrachloro-m-xylene	4.334	4.236	4.436	54.110	50.000	8.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/04/2025 12/04/2025

Client Sample No.: CCAL04 **Date Analyzed:** 12/05/2025
Lab Sample No.: AR1242CCC500 **Data File :** PO115625.D **Time Analyzed:** 15:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1242-1	4.712	4.615	4.815	545.600	500.000	9.1
Aroclor-1242-2	4.731	4.634	4.834	523.100	500.000	4.6
Aroclor-1242-3	4.905	4.808	5.008	530.190	500.000	6.0
Aroclor-1242-4	4.989	4.892	5.092	531.250	500.000	6.3
Aroclor-1242-5	5.508	5.412	5.612	562.670	500.000	12.5
Decachlorobiphenyl	8.618	8.522	8.722	49.870	50.000	-0.3
Tetrachloro-m-xylene	3.637	3.538	3.738	50.450	50.000	0.9