

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ALLI03  
**Lab Code:** ACE **SDG NO.:** Q3483  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/04/2025 12/04/2025

**Client Sample No.:** CCAL05 **Date Analyzed:** 12/08/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO115631.D **Time Analyzed:** 09:58

| COMPOUND             | RT    | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|--------|--------------------|-------------------|------|
|                      |       | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.479 | 5.375     | 5.575  | 512.430            | 500.000           | 2.5  |
| Aroclor-1016-2       | 5.501 | 5.396     | 5.596  | 525.930            | 500.000           | 5.2  |
| Aroclor-1016-3       | 5.562 | 5.458     | 5.658  | 529.650            | 500.000           | 5.9  |
| Aroclor-1016-4       | 5.658 | 5.553     | 5.753  | 531.530            | 500.000           | 6.3  |
| Aroclor-1016-5       | 5.949 | 5.846     | 6.046  | 518.500            | 500.000           | 3.7  |
| Aroclor-1260-1       | 7.062 | 6.958     | 7.158  | 533.850            | 500.000           | 6.8  |
| Aroclor-1260-2       | 7.317 | 7.212     | 7.412  | 496.270            | 500.000           | -0.7 |
| Aroclor-1260-3       | 7.674 | 7.570     | 7.770  | 540.430            | 500.000           | 8.1  |
| Aroclor-1260-4       | 7.896 | 7.791     | 7.991  | 530.300            | 500.000           | 6.1  |
| Aroclor-1260-5       | 8.202 | 8.098     | 8.298  | 523.810            | 500.000           | 4.8  |
| Decachlorobiphenyl   | 9.919 | 9.812     | 10.012 | 52.060             | 50.000            | 4.1  |
| Tetrachloro-m-xylene | 4.340 | 4.236     | 4.436  | 51.650             | 50.000            | 3.3  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ALLI03  
**Lab Code:** ACE **SDG NO.:** Q3483  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/04/2025 12/04/2025

**Client Sample No.:** CCAL05 **Date Analyzed:** 12/08/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO115631.D **Time Analyzed:** 09:58

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.714 | 4.615     | 4.815 | 485.740            | 500.000           | -2.9 |
| Aroclor-1016-2       | 4.733 | 4.634     | 4.834 | 482.870            | 500.000           | -3.4 |
| Aroclor-1016-3       | 4.907 | 4.808     | 5.008 | 473.370            | 500.000           | -5.3 |
| Aroclor-1016-4       | 4.950 | 4.851     | 5.051 | 502.470            | 500.000           | 0.5  |
| Aroclor-1016-5       | 5.162 | 5.062     | 5.262 | 495.520            | 500.000           | -0.9 |
| Aroclor-1260-1       | 6.190 | 6.091     | 6.291 | 487.770            | 500.000           | -2.4 |
| Aroclor-1260-2       | 6.378 | 6.278     | 6.478 | 481.680            | 500.000           | -3.7 |
| Aroclor-1260-3       | 6.530 | 6.431     | 6.631 | 456.900            | 500.000           | -8.6 |
| Aroclor-1260-4       | 7.001 | 6.901     | 7.101 | 493.370            | 500.000           | -1.3 |
| Aroclor-1260-5       | 7.242 | 7.142     | 7.342 | 489.380            | 500.000           | -2.1 |
| Decachlorobiphenyl   | 8.621 | 8.522     | 8.722 | 48.550             | 50.000            | -2.9 |
| Tetrachloro-m-xylene | 3.638 | 3.538     | 3.738 | 47.940             | 50.000            | -4.1 |