

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/04/2025 12/04/2025

Client Sample No.: CCAL06 **Date Analyzed:** 12/08/2025
Lab Sample No.: AR1254CCC500 **Data File :** PO115634.D **Time Analyzed:** 10:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	6.315	6.216	6.416	469.740	500.000	-6.1
Aroclor-1254-2	6.532	6.431	6.631	472.060	500.000	-5.6
Aroclor-1254-3	6.891	6.792	6.992	470.730	500.000	-5.9
Aroclor-1254-4	7.174	7.073	7.273	476.920	500.000	-4.6
Aroclor-1254-5	7.589	7.489	7.689	475.850	500.000	-4.8
Decachlorobiphenyl	9.907	9.809	10.009	48.450	50.000	-3.1
Tetrachloro-m-xylene	4.335	4.233	4.433	46.640	50.000	-6.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/04/2025 12/04/2025

Client Sample No.: CCAL06 **Date Analyzed:** 12/08/2025
Lab Sample No.: AR1254CCC500 **Data File :** PO115634.D **Time Analyzed:** 10:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	5.509	5.409	5.609	453.280	500.000	-9.3
Aroclor-1254-2	5.656	5.557	5.757	457.710	500.000	-8.5
Aroclor-1254-3	6.058	5.958	6.158	462.620	500.000	-7.5
Aroclor-1254-4	6.286	6.186	6.386	465.780	500.000	-6.8
Aroclor-1254-5	6.702	6.602	6.802	464.970	500.000	-7.0
Decachlorobiphenyl	8.619	8.520	8.720	44.690	50.000	-10.6
Tetrachloro-m-xylene	3.637	3.536	3.736	46.410	50.000	-7.2