

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/04/2025 12/04/2025

Client Sample No.: CCAL07 **Date Analyzed:** 12/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115640.D **Time Analyzed:** 13:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.474	5.375	5.575	503.630	500.000	0.7
Aroclor-1016-2	5.495	5.396	5.596	527.490	500.000	5.5
Aroclor-1016-3	5.556	5.458	5.658	528.120	500.000	5.6
Aroclor-1016-4	5.652	5.553	5.753	530.660	500.000	6.1
Aroclor-1016-5	5.944	5.846	6.046	519.800	500.000	4.0
Aroclor-1260-1	7.057	6.958	7.158	538.050	500.000	7.6
Aroclor-1260-2	7.311	7.212	7.412	487.740	500.000	-2.5
Aroclor-1260-3	7.668	7.570	7.770	545.130	500.000	9.0
Aroclor-1260-4	7.890	7.791	7.991	530.280	500.000	6.1
Aroclor-1260-5	8.196	8.098	8.298	531.690	500.000	6.3
Decachlorobiphenyl	9.909	9.812	10.012	53.280	50.000	6.6
Tetrachloro-m-xylene	4.335	4.236	4.436	52.290	50.000	4.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/04/2025 12/04/2025

Client Sample No.: CCAL07 **Date Analyzed:** 12/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115640.D **Time Analyzed:** 13:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.713	4.615	4.815	492.250	500.000	-1.6
Aroclor-1016-2	4.732	4.634	4.834	499.260	500.000	-0.1
Aroclor-1016-3	4.906	4.808	5.008	482.120	500.000	-3.6
Aroclor-1016-4	4.948	4.851	5.051	513.340	500.000	2.7
Aroclor-1016-5	5.160	5.062	5.262	506.780	500.000	1.4
Aroclor-1260-1	6.188	6.091	6.291	498.020	500.000	-0.4
Aroclor-1260-2	6.376	6.278	6.478	490.060	500.000	-2.0
Aroclor-1260-3	6.528	6.431	6.631	467.200	500.000	-6.6
Aroclor-1260-4	6.999	6.901	7.101	499.400	500.000	-0.1
Aroclor-1260-5	7.240	7.142	7.342	497.660	500.000	-0.5
Decachlorobiphenyl	8.618	8.522	8.722	49.970	50.000	-0.1
Tetrachloro-m-xylene	3.637	3.538	3.738	49.310	50.000	-1.4