

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL10 **Date Analyzed:** 12/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076815.D **Time Analyzed:** 19:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.773	5.667	5.867	431.240	500.000	-13.8
Aroclor-1016-2	5.794	5.687	5.887	433.790	500.000	-13.2
Aroclor-1016-3	5.856	5.750	5.950	429.390	500.000	-14.1
Aroclor-1016-4	5.953	5.847	6.047	430.040	500.000	-14.0
Aroclor-1016-5	6.245	6.140	6.340	435.480	500.000	-12.9
Aroclor-1260-1	7.362	7.256	7.456	430.590	500.000	-13.9
Aroclor-1260-2	7.615	7.509	7.709	419.100	500.000	-16.2
Aroclor-1260-3	7.972	7.867	8.067	422.450	500.000	-15.5
Aroclor-1260-4	8.198	8.093	8.293	427.500	500.000	-14.5
Aroclor-1260-5	8.522	8.416	8.616	417.810	500.000	-16.4
Decachlorobiphenyl	10.368	10.265	10.465	42.190	50.000	-15.6
Tetrachloro-m-xylene	4.621	4.514	4.714	43.510	50.000	-13.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL10 **Date Analyzed:** 12/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076815.D **Time Analyzed:** 19:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.852	4.750	4.950	480.670	500.000	-3.9
Aroclor-1016-2	4.869	4.767	4.967	493.250	500.000	-1.4
Aroclor-1016-3	5.044	4.942	5.142	490.370	500.000	-1.9
Aroclor-1016-4	5.087	4.984	5.184	453.630	500.000	-9.3
Aroclor-1016-5	5.299	5.197	5.397	538.650	500.000	7.7
Aroclor-1260-1	6.326	6.224	6.424	489.980	500.000	-2.0
Aroclor-1260-2	6.513	6.412	6.612	494.910	500.000	-1.0
Aroclor-1260-3	6.665	6.564	6.764	470.100	500.000	-6.0
Aroclor-1260-4	7.134	7.033	7.233	476.940	500.000	-4.6
Aroclor-1260-5	7.375	7.273	7.473	469.470	500.000	-6.1
Decachlorobiphenyl	8.766	8.666	8.866	45.350	50.000	-9.3
Tetrachloro-m-xylene	3.774	3.671	3.871	49.580	50.000	-0.8