

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ALLI03  
**Lab Code:** ACE **SDG NO.:** Q3483  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

**Client Sample No.:** CCAL11 **Date Analyzed:** 11/19/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PQ071332.D **Time Analyzed:** 13:04

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.497 | 4.397     | 4.597 | 481.490            | 500.000           | -3.7 |
| Aroclor-1016-2       | 4.516 | 4.416     | 4.616 | 488.020            | 500.000           | -2.4 |
| Aroclor-1016-3       | 4.573 | 4.474     | 4.674 | 485.270            | 500.000           | -2.9 |
| Aroclor-1016-4       | 4.664 | 4.564     | 4.764 | 487.200            | 500.000           | -2.6 |
| Aroclor-1016-5       | 4.949 | 4.849     | 5.049 | 488.530            | 500.000           | -2.3 |
| Aroclor-1260-1       | 6.042 | 5.943     | 6.143 | 464.240            | 500.000           | -7.2 |
| Aroclor-1260-2       | 6.299 | 6.200     | 6.400 | 486.320            | 500.000           | -2.7 |
| Aroclor-1260-3       | 6.652 | 6.552     | 6.752 | 484.460            | 500.000           | -3.1 |
| Aroclor-1260-4       | 6.867 | 6.767     | 6.967 | 489.140            | 500.000           | -2.2 |
| Aroclor-1260-5       | 7.174 | 7.075     | 7.275 | 494.950            | 500.000           | -1.0 |
| Decachlorobiphenyl   | 8.536 | 8.437     | 8.637 | 49.110             | 50.000            | -1.8 |
| Tetrachloro-m-xylene | 3.411 | 3.311     | 3.511 | 49.280             | 50.000            | -1.4 |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Contract: ALLI03

Lab Code: ACE SDG NO.: Q3483

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2025 11/18/2025

Client Sample No.: CCAL11 Date Analyzed: 11/19/2025

Lab Sample No.: AR1660CCC500 Data File : PQ071332.D Time Analyzed: 13:04

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Aroclor-1016-1       | 3.870 | 3.771                | 3.971 | 537.530            | 500.000           | 7.5  |
| Aroclor-1016-2       | 3.886 | 3.788                | 3.988 | 553.750            | 500.000           | 10.8 |
| Aroclor-1016-3       | 4.048 | 3.950                | 4.150 | 550.090            | 500.000           | 10.0 |
| Aroclor-1016-4       | 4.097 | 3.998                | 4.198 | 536.250            | 500.000           | 7.3  |
| Aroclor-1016-5       | 4.294 | 4.195                | 4.395 | 544.870            | 500.000           | 9.0  |
| Aroclor-1260-1       | 5.295 | 5.197                | 5.397 | 539.360            | 500.000           | 7.9  |
| Aroclor-1260-2       | 5.484 | 5.386                | 5.586 | 547.740            | 500.000           | 9.5  |
| Aroclor-1260-3       | 5.628 | 5.530                | 5.730 | 539.920            | 500.000           | 8.0  |
| Aroclor-1260-4       | 6.092 | 5.993                | 6.193 | 544.730            | 500.000           | 8.9  |
| Aroclor-1260-5       | 6.336 | 6.238                | 6.438 | 543.770            | 500.000           | 8.8  |
| Decachlorobiphenyl   | 7.657 | 7.559                | 7.759 | 54.640             | 50.000            | 9.3  |
| Tetrachloro-m-xylene | 2.855 | 2.756                | 2.956 | 54.730             | 50.000            | 9.5  |