

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: ALLI03
SDG NO.: Q3483

Calibration Date(s): 11/14/2025 11/14/2025
Calibration Times: 20:56 21:51

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD091196.D</u>	CF 075 =	<u>PD091197.D</u>		
CF 050 =		<u>PD091198.D</u>	CF 025 =	<u>PD091199.D</u>	CF 005 =	<u>PD091200.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	4592720000	4750410000	4706670000	4762630000	5745110000	4911510000	10
4,4'-DDE	5537410000	5724420000	5594750000	5596930000	6543500000	5799400000	7
4,4'-DDT	4033900000	4103640000	4023370000	3973800000	4542840000	4135510000	6
Aldrin	6913630000	7108360000	6994820000	7033330000	8231840000	7256400000	8
alpha-BHC	7991490000	8120730000	7870330000	7613270000	8381750000	7995510000	4
alpha-Chlordane	6168980000	6320470000	6262210000	6530520000	9131880000	6882810000	18
beta-BHC	2584470000	2701070000	2721620000	2850470000	3669860000	2905500000	15
Decachlorobiphenyl	4266550000	4520270000	4633000000	4951710000	6537810000	4981870000	18
delta-BHC	7343040000	7477970000	7279360000	7082560000	7977500000	7432080000	5
Dieldrin	6112350000	6323390000	6220240000	6264150000	7404460000	6464920000	8
Endosulfan I	5622190000	5847020000	5809210000	5991120000	7445430000	6142990000	12
Endosulfan II	5022110000	5221970000	5253890000	5419940000	6878740000	5559330000	14
Endosulfan sulfate	4635330000	4824080000	4867360000	5023730000	6373440000	5144790000	14
Endrin	4763010000	4861760000	4775160000	4762720000	5654790000	4963490000	8
Endrin aldehyde	3642590000	3820110000	3882150000	4087850000	5397480000	4166040000	17
Endrin ketone	5131250000	5317460000	5324710000	5507230000	6887120000	5633560000	13
gamma-BHC (Lindane)	7349570000	7492470000	7318000000	7162850000	8041460000	7472870000	5
gamma-Chlordane	6123460000	6343120000	6211230000	6313670000	7512390000	6500770000	9
Heptachlor	6873030000	7051630000	6922840000	6902860000	8058350000	7161740000	7
Heptachlor epoxide	6045880000	6198590000	6197710000	6327850000	7713630000	6496730000	11
Methoxychlor	1967810000	2033910000	2048060000	2143440000	2678460000	2174340000	13
Mirex	3028840000	3191280000	3284080000	3546030000	4734970000	3557040000	19
Tetrachloro-m-xylene	3100550000	3224630000	3262160000	3454950000	4200860000	3448630000	13

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: ALLI03
SDG NO.: Q3483

Calibration Date(s): 11/14/2025 11/14/2025
Calibration Times: 20:56 21:51

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD091196.D</u>	CF 075 =	<u>PD091197.D</u>		
CF 050 =	<u>PD091198.D</u>	CF 025 =	<u>PD091199.D</u>	CF 005 =	<u>PD091200.D</u>		
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	42169800000	44652200000	46098100000	48914200000	62508300000	48868500000	16
4,4'-DDE	46784900000	49153600000	50413800000	53041500000	67816900000	53442100000	16
4,4'-DDT	36565400000	37844700000	38225900000	38748200000	44278500000	39132500000	8
Aldrin	49756900000	52224800000	53642000000	56183400000	70485000000	56458400000	14
alpha-BHC	57081100000	59401300000	60005600000	62092800000	77433200000	63202800000	13
alpha-Chlordane	46162300000	48496400000	49694800000	52042000000	65261400000	52331400000	14
beta-BHC	20855800000	21963600000	22629700000	23880400000	30805800000	24027100000	16
Decachlorobiphenyl	35391500000	36530600000	37125400000	38755300000	50071600000	39574900000	15
delta-BHC	51951800000	54520500000	55333700000	57523500000	73056200000	58477100000	14
Dieldrin	47342900000	49974500000	51581700000	54390600000	68265900000	54311100000	15
Endosulfan I	41219300000	43810000000	45334200000	48114600000	60781000000	47851800000	16
Endosulfan II	40233600000	42630400000	44155200000	46739900000	59639900000	46679800000	16
Endosulfan sulfate	38807500000	40991800000	42555300000	44718700000	57101400000	44834900000	16
Endrin	39723100000	41747800000	43010100000	44994600000	56487600000	45192700000	15
Endrin aldehyde	29904600000	31738200000	33159100000	35293100000	46350200000	35289100000	18
Endrin ketone	43743400000	46381200000	48213200000	51142400000	65299300000	50955900000	17
gamma-BHC (Lindane)	52036700000	54159800000	54955200000	56993600000	70961400000	57821300000	13
gamma-Chlordane	48541600000	50822000000	51970500000	54222400000	67580500000	54627400000	14
Heptachlor	48452800000	51096600000	52341200000	55061200000	69095300000	55209400000	15
Heptachlor epoxide	43794700000	46344600000	48042100000	50883900000	65728100000	50958700000	17
Methoxychlor	17875100000	18763500000	19368600000	20213700000	24036200000	20051400000	12
Mirex	28591700000	30079900000	31140600000	33169500000	42718000000	33140000000	17
Tetrachloro-m-xylene	32274400000	33875900000	34577100000	36418900000	46004800000	36630200000	15