

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091194.D Date Analyzed: 11/14/2025

Lab Sample No.(PEM): PEM Time Analyzed: 20:15

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.064	8.960	9.160	20.170	20.000	0.9
Tetrachloro-m-xylene	3.544	3.490	3.590	19.500	20.000	-2.5
alpha-BHC	3.993	3.940	4.040	8.250	10.000	-17.5
beta-BHC	4.511	4.460	4.560	9.240	10.000	-7.6
gamma-BHC (Lindane)	4.324	4.270	4.370	8.450	10.000	-15.5
Endrin	6.567	6.500	6.640	44.390	50.000	-11.2
4,4'-DDT	7.015	6.940	7.090	88.290	100.000	-11.7
Methoxychlor	7.487	7.420	7.560	207.040	250.000	-17.2

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091194.D Date Analyzed: 11/14/2025

Lab Sample No.(PEM): PEM Time Analyzed: 20:15

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.059	7.960	8.160	19.620	20.000	-1.9
Tetrachloro-m-xylene	2.874	2.820	2.920	19.370	20.000	-3.2
alpha-BHC	3.385	3.330	3.440	9.380	10.000	-6.2
beta-BHC	4.018	3.970	4.070	9.530	10.000	-4.7
gamma-BHC (Lindane)	3.721	3.670	3.770	9.460	10.000	-5.4
Endrin	5.779	5.710	5.850	45.450	50.000	-9.1
4,4'-DDT	6.172	6.100	6.240	89.340	100.000	-10.7
Methoxychlor	6.742	6.670	6.810	186.630	250.000	-25.3