

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111525

Review By	rahul	Review On	11/14/2025 3:12:58 PM
Supervise By	Sohil	Supervise On	11/17/2025 2:14:18 PM
SubDirectory	PD111525	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP25075		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	I.BLK	I.BLK	PD091193.D	14 Nov 2025 20:01		AR\AJ	Ok
2	PEM	PEM	PD091194.D	14 Nov 2025 20:15		AR\AJ	Ok,M
3	RESCHK	RESCHK	PD091195.D	14 Nov 2025 20:43		AR\AJ	Ok
4	PSTDICC100	PSTDICC100	PD091196.D	14 Nov 2025 20:56		AR\AJ	Ok
5	PSTDICC075	PSTDICC075	PD091197.D	14 Nov 2025 21:10		AR\AJ	Ok
6	PSTDICC050	PSTDICC050	PD091198.D	14 Nov 2025 21:24		AR\AJ	Ok
7	PSTDICC025	PSTDICC025	PD091199.D	14 Nov 2025 21:37		AR\AJ	Ok
8	PSTDICC005	PSTDICC005	PD091200.D	14 Nov 2025 21:51		AR\AJ	Ok,M
9	PCHLORICC1000	PCHLORICC1000	PD091201.D	14 Nov 2025 22:04		AR\AJ	Ok
10	PCHLORICC750	PCHLORICC750	PD091202.D	14 Nov 2025 22:18		AR\AJ	Ok
11	PCHLORICC500	PCHLORICC500	PD091203.D	14 Nov 2025 22:32		AR\AJ	Ok
12	PCHLORICC250	PCHLORICC250	PD091204.D	14 Nov 2025 22:46		AR\AJ	Ok,M
13	PCHLORICC50	PCHLORICC50	PD091205.D	14 Nov 2025 22:59		AR\AJ	Ok,M
14	PTOXICC1000	PTOXICC1000	PD091206.D	14 Nov 2025 23:13		AR\AJ	Ok
15	PTOXICC750	PTOXICC750	PD091207.D	14 Nov 2025 23:27		AR\AJ	Ok
16	PTOXICC500	PTOXICC500	PD091208.D	14 Nov 2025 23:40		AR\AJ	Ok
17	PTOXICC250	PTOXICC250	PD091209.D	14 Nov 2025 23:54		AR\AJ	Ok
18	PTOXICC100	PTOXICC100	PD091210.D	15 Nov 2025 00:08		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111525

Review By	rahul	Review On	11/14/2025 3:12:58 PM
Supervise By	Sohil	Supervise On	11/17/2025 2:14:18 PM
SubDirectory	PD111525	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP25075		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PSTDICV050	ICVPD111525	PD091211.D	15 Nov 2025 00:21		AR\AJ	Ok
20	PCHLORICV050	ICVPD111525CHLOR	PD091212.D	15 Nov 2025 00:49		AR\AJ	Ok
21	PTOXICV050	ICVPD111525TOX	PD091213.D	15 Nov 2025 01:30		AR\AJ	Ok
22	I.BLK	I.BLK	PD091214.D	15 Nov 2025 01:44		AR\AJ	Ok
23	PEM	PEM	PD091215.D	15 Nov 2025 01:57		AR\AJ	Ok,M
24	PSTDCCC050	PSTDCCC050	PD091216.D	15 Nov 2025 02:11		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111925

Review By	Abdul	Review On	11/20/2025 8:20:36 AM
Supervise By	Sohil	Supervise On	11/20/2025 4:03:21 PM
SubDirectory	PD111925	HP Acquire Method	HP Processing Method pd111525 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24942,PP25075
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD091263.D	19 Nov 2025 08:52		AR\AJ	Ok
2	I.BLK	I.BLK	PD091264.D	19 Nov 2025 09:05		AR\AJ	Ok
3	PEM	PEM	PD091265.D	19 Nov 2025 09:19		AR\AJ	Ok
4	PSTDCCC050	PSTDCCC050	PD091266.D	19 Nov 2025 12:24		AR\AJ	Ok
5	Q3660-01	FRAC-TANK-M21012	PD091267.D	19 Nov 2025 12:38		AR\AJ	Ok
6	PB170610BS	PB170610BS	PD091268.D	19 Nov 2025 13:33	TCMX high in both column	AR\AJ	Not Ok
7	PB170610BSD	PB170610BSD	PD091269.D	19 Nov 2025 13:54	TCMX high in both column	AR\AJ	Not Ok
8	PB170628BL	PB170628BL	PD091270.D	19 Nov 2025 14:07		AR\AJ	Ok
9	PB170628BS	PB170628BS	PD091271.D	19 Nov 2025 14:21	DCB high in 2nd column	AR\AJ	Not Ok
10	Q3483-20	HW1025-PT-PEST-SO	PD091272.D	19 Nov 2025 14:34	need dilution	AR\AJ	Dilution
11	Q3669-01	EO-01-11182025	PD091273.D	19 Nov 2025 14:48		AR\AJ	Ok,M
12	Q3483-20DL	HW1025-PT-PEST-SO	PD091274.D	19 Nov 2025 15:02	TCMX high in 2nd column	AR\AJ	Ok,M
13	I.BLK	I.BLK	PD091275.D	19 Nov 2025 15:16		AR\AJ	Ok
14	PSTDCCC050	PSTDCCC050	PD091276.D	19 Nov 2025 15:55		AR\AJ	Ok
15	PP25072	PP25072	PD091277.D	19 Nov 2025 16:09		AR\AJ	Ok
16	Q3662-02	B3-0.0-1.0-20251117	PD091278.D	19 Nov 2025 16:24		AR\AJ	Ok,M
17	Q3662-02MS	B3-0.0-1.0-20251117MS	PD091279.D	19 Nov 2025 16:37		AR\AJ	Ok,M
18	Q3662-02MSD	B3-0.0-1.0-20251117MSD	PD091280.D	19 Nov 2025 16:51	RPD Fail	AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111925

Review By	Abdul	Review On	11/20/2025 8:20:36 AM
Supervise By	Sohil	Supervise On	11/20/2025 4:03:21 PM
SubDirectory	PD111925	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP25075		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q3662-04	B1-0.0-1.0-20251117	PD091281.D	19 Nov 2025 17:05		AR\AJ	Ok,M
20	Q3664-01	ARS20-0001	PD091282.D	19 Nov 2025 17:18		AR\AJ	Ok,M
21	Q3664-03	ARS20-0013	PD091283.D	19 Nov 2025 17:32		AR\AJ	Ok,M
22	Q3664-05	291399	PD091284.D	19 Nov 2025 17:45		AR\AJ	Ok,M
23	I.BLK	I.BLK	PD091285.D	19 Nov 2025 17:59		AR\AJ	Ok
24	PEM	PEM	PD091286.D	19 Nov 2025 18:13		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PD091287.D	19 Nov 2025 18:40		AR\AJ	Ok,M
26	PB170638BL	PB170638BL	PD091288.D	19 Nov 2025 18:54		AR\AJ	Ok
27	PB170638BS	PB170638BS	PD091289.D	19 Nov 2025 19:07	some compound recovery fail	AR\AJ	Not Ok
28	PB170600TB	PB170600TB	PD091290.D	19 Nov 2025 19:21		AR\AJ	Ok
29	Q3659-01	TANK-1-COMP-1	PD091291.D	19 Nov 2025 19:35		AR\AJ	Ok
30	Q3659-01MS	TANK-1-COMP-1MS	PD091292.D	19 Nov 2025 19:48		AR\AJ	Ok,M
31	Q3659-01MSD	TANK-1-COMP-1MSD	PD091293.D	19 Nov 2025 20:02		AR\AJ	Ok,M
32	Q3659-02	TANK-1-COMP-2	PD091294.D	19 Nov 2025 20:16		AR\AJ	Ok
33	I.BLK	I.BLK	PD091295.D	19 Nov 2025 20:29		AR\AJ	Ok
34	PSTDCCC050	PSTDCCC050	PD091296.D	19 Nov 2025 20:43		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD120125

Review By	Abdul	Review On	12/2/2025 8:48:11 AM
Supervise By	mohammad	Supervise On	12/3/2025 12:24:21 AM
SubDirectory	PD120125	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP25075		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD091400.D	01 Dec 2025 09:45		AR\AJ	Ok
2	I.BLK	I.BLK	PD091401.D	01 Dec 2025 09:59		AR\AJ	Ok
3	PEM	PEM	PD091402.D	01 Dec 2025 10:12		AR\AJ	Ok
4	PSTDCCC050	PSTDCCC050	PD091403.D	01 Dec 2025 10:26		AR\AJ	Ok
5	PB170628BS	PB170628BS	PD091404.D	01 Dec 2025 12:10		AR\AJ	Ok
6	PB170762BL	PB170762BL	PD091405.D	01 Dec 2025 12:32		AR\AJ	Ok
7	PB170762BS	PB170762BS	PD091406.D	01 Dec 2025 12:45		AR\AJ	Ok
8	Q3725-04	STOCKPILE-SAMPLES	PD091407.D	01 Dec 2025 12:59		AR\AJ	Ok,M
9	Q3732-01	HD-01-11-26-2025	PD091408.D	01 Dec 2025 13:12		AR\AJ	Ok,M
10	Q3733-01	SU-04-11-26-2025	PD091409.D	01 Dec 2025 13:26		AR\AJ	Ok,M
11	Q3735-01	BU-3-112625	PD091410.D	01 Dec 2025 13:40	Tcmx high in 1st column	AR\AJ	Ok,M
12	Q3735-01MS	BU-3-112625MS	PD091411.D	01 Dec 2025 13:53	Comp # 12 fail for recovery	AR\AJ	Ok,M
13	Q3735-01MSD	BU-3-112625MSD	PD091412.D	01 Dec 2025 14:07	Comp # 12,14 fail for recovery	AR\AJ	Ok,M
14	I.BLK	I.BLK	PD091413.D	01 Dec 2025 14:21		AR\AJ	Ok
15	PSTDCCC050	PSTDCCC050	PD091414.D	01 Dec 2025 15:11		AR\AJ	Ok,M

M : Manual Integration