

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Continuing Calib Date: 11/20/2025

Initial Calibration Date(s): 11/14/2025

11/14/2025

Continuing Calib Time: 13:25

Initial Calibration Time(s): 22:04

22:59

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Chlordane-1 (1)	4.70	4.71	4.61	4.81	0.01
Chlordane-2 (2)	5.23	5.23	5.13	5.33	0.00
Chlordane-3 (3)	5.94	5.94	5.84	6.04	0.00
Chlordane-4 (4)	6.02	6.02	5.92	6.12	0.00
Chlordane-5 (5)	6.86	6.86	6.76	6.96	0.00
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Continuing Calib Date: 11/20/2025

Initial Calibration Date(s): 11/14/2025

11/14/2025

Continuing Calib Time: 13:25

Initial Calibration Time(s): 22:04

22:59

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	3.89	3.90	3.80	4.00	0.01
Chlordane-2	(2)	4.47	4.48	4.38	4.58	0.01
Chlordane-3	(3)	5.11	5.11	5.01	5.21	0.00
Chlordane-4	(4)	5.18	5.18	5.08	5.28	0.00
Chlordane-5	(5)	6.08	6.08	5.98	6.18	0.01
Decachlorobiphenyl		8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene		2.87	2.87	2.77	2.97	0.00