

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/20/2025
Lab Sample No.: PCHLORCCC500 **Data File :** PD091301.D **Time Analyzed:** 09:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	4.712	4.607	4.807	459.630	500.000	-8.1
Chlordane-2	5.239	5.133	5.333	443.710	500.000	-11.3
Chlordane-3	5.944	5.839	6.039	461.190	500.000	-7.8
Chlordane-4	6.029	5.924	6.124	443.480	500.000	-11.3
Chlordane-5	6.868	6.764	6.964	440.680	500.000	-11.9
Decachlorobiphenyl	9.069	8.963	9.163	40.440	50.000	-19.1
Tetrachloro-m-xylene	3.549	3.443	3.643	41.330	50.000	-17.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/20/2025
Lab Sample No.: PCHLORCCC500 **Data File :** PD091301.D **Time Analyzed:** 09:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	3.894	3.796	3.996	475.640	500.000	-4.9
Chlordane-2	4.474	4.376	4.576	475.980	500.000	-4.8
Chlordane-3	5.113	5.014	5.214	479.060	500.000	-4.2
Chlordane-4	5.177	5.078	5.278	480.000	500.000	-4.0
Chlordane-5	6.077	5.978	6.178	489.520	500.000	-2.1
Decachlorobiphenyl	8.056	7.958	8.158	45.360	50.000	-9.3
Tetrachloro-m-xylene	2.870	2.773	2.973	53.570	50.000	7.1