

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/20/2025
Lab Sample No.: PCHLORCCC500 **Data File :** PD091308.D **Time Analyzed:** 13:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	4.704	4.607	4.807	460.620	500.000	-7.9
Chlordane-2	5.231	5.133	5.333	447.270	500.000	-10.5
Chlordane-3	5.937	5.839	6.039	461.470	500.000	-7.7
Chlordane-4	6.021	5.924	6.124	447.800	500.000	-10.4
Chlordane-5	6.861	6.764	6.964	445.520	500.000	-10.9
Decachlorobiphenyl	9.059	8.963	9.163	41.010	50.000	-18.0
Tetrachloro-m-xylene	3.542	3.443	3.643	41.750	50.000	-16.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/20/2025
Lab Sample No.: PCHLORCCC500 **Data File :** PD091308.D **Time Analyzed:** 13:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	3.893	3.796	3.996	488.960	500.000	-2.2
Chlordane-2	4.474	4.376	4.576	489.030	500.000	-2.2
Chlordane-3	5.112	5.014	5.214	489.860	500.000	-2.0
Chlordane-4	5.176	5.078	5.278	489.360	500.000	-2.1
Chlordane-5	6.075	5.978	6.178	498.070	500.000	-0.4
Decachlorobiphenyl	8.055	7.958	8.158	46.190	50.000	-7.6
Tetrachloro-m-xylene	2.871	2.773	2.973	54.130	50.000	8.3