

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Continuing Calib Date: 11/20/2025

Initial Calibration Date(s): 11/14/2025

11/15/2025

Continuing Calib Time: 14:09

Initial Calibration Time(s): 23:13

00:08

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Toxaphene-1	(1)	6.23	6.23	6.13	6.33	0.00
Toxaphene-2	(2)	6.43	6.44	6.34	6.54	0.01
Toxaphene-3	(3)	7.14	7.14	7.04	7.24	0.00
Toxaphene-4	(4)	7.55	7.56	7.46	7.66	0.01
Toxaphene-5	(5)	7.92	7.92	7.82	8.02	0.00
Decachlorobiphenyl		9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene		3.54	3.54	3.44	3.64	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3483

Continuing Calib Date: 11/20/2025

Initial Calibration Date(s): 11/14/2025

11/15/2025

Continuing Calib Time: 14:09

Initial Calibration Time(s): 23:13

00:08

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Toxaphene-1	(1)	5.46	5.46	5.36	5.56	0.00
Toxaphene-2	(2)	5.63	5.64	5.54	5.74	0.01
Toxaphene-3	(3)	6.74	6.74	6.64	6.84	0.00
Toxaphene-4	(4)	7.18	7.19	7.09	7.29	0.01
Toxaphene-5	(5)	7.31	7.32	7.22	7.42	0.01
Decachlorobiphenyl		8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene		2.87	2.87	2.77	2.97	0.00