

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL03 **Date Analyzed:** 12/04/2025
Lab Sample No.: PTOXCCC500 **Data File :** PD091451.D **Time Analyzed:** 12:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.069	8.963	9.163	48.730	50.000	-2.5
Tetrachloro-m-xylene	3.547	3.443	3.643	51.740	50.000	3.5
Toxaphene-1	6.238	6.134	6.334	403.220	500.000	-19.4
Toxaphene-2	6.440	6.335	6.535	483.000	500.000	-3.4
Toxaphene-3	7.145	7.041	7.241	482.200	500.000	-3.6
Toxaphene-4	7.561	7.456	7.656	481.820	500.000	-3.6
Toxaphene-5	7.927	7.822	8.022	483.520	500.000	-3.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL03 **Date Analyzed:** 12/04/2025
Lab Sample No.: PTOXCCC500 **Data File :** PD091451.D **Time Analyzed:** 12:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.057	7.958	8.158	56.010	50.000	12.0
Tetrachloro-m-xylene	2.871	2.774	2.974	58.600	50.000	17.2
Toxaphene-1	5.463	5.363	5.563	539.770	500.000	8.0
Toxaphene-2	5.634	5.535	5.735	557.700	500.000	11.5
Toxaphene-3	6.744	6.643	6.843	555.350	500.000	11.1
Toxaphene-4	7.185	7.086	7.286	563.540	500.000	12.7
Toxaphene-5	7.316	7.217	7.417	532.920	500.000	6.6