

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL04 **Date Analyzed:** 12/04/2025
Lab Sample No.: PTOXCCC500 **Data File :** PD091455.D **Time Analyzed:** 14:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.069	8.963	9.163	45.770	50.000	-8.5
Tetrachloro-m-xylene	3.548	3.443	3.643	47.820	50.000	-4.4
Toxaphene-1	6.240	6.134	6.334	535.910	500.000	7.2
Toxaphene-2	6.440	6.335	6.535	451.760	500.000	-9.6
Toxaphene-3	7.146	7.041	7.241	445.080	500.000	-11.0
Toxaphene-4	7.562	7.456	7.656	445.150	500.000	-11.0
Toxaphene-5	7.927	7.822	8.022	447.240	500.000	-10.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3483
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL04 **Date Analyzed:** 12/04/2025
Lab Sample No.: PTOXCCC500 **Data File :** PD091455.D **Time Analyzed:** 14:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.057	7.958	8.158	51.700	50.000	3.4
Tetrachloro-m-xylene	2.871	2.774	2.974	54.220	50.000	8.4
Toxaphene-1	5.463	5.363	5.563	539.950	500.000	8.0
Toxaphene-2	5.635	5.535	5.735	516.850	500.000	3.4
Toxaphene-3	6.743	6.643	6.843	509.370	500.000	1.9
Toxaphene-4	7.186	7.086	7.286	527.340	500.000	5.5
Toxaphene-5	7.317	7.217	7.417	491.000	500.000	-1.8