

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112725\
 Data File : BP026190.D
 Acq On : 26 Nov 2025 14:04
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 14:52:55 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
2	1,4-Dioxane	0.505	0.486	3.8	78	0.00
3	Pyridine	1.461	1.358	7.0	73	0.00
4	n-Nitrosodimethylamine	0.546	0.500	8.4	72	0.00
5 S	2-Fluorophenol	1.246	1.185	4.9	74	0.00
6	Aniline	2.094	1.972	5.8	73	0.00
7 S	Phenol-d6	1.586	1.536	3.2	74	-0.01
8	2-Chlorophenol	1.418	1.403	1.1	77	0.00
9	Benzaldehyde	0.831	0.859	-3.4	73	0.00
10 C	Phenol	1.709	1.640	4.0	73	0.00
11	bis(2-Chloroethyl)ether	1.338	1.296	3.1	74	0.00
12	1,3-Dichlorobenzene	1.528	1.553	-1.6	80	-0.01
13 C	1,4-Dichlorobenzene	1.540	1.553	-0.8	79	0.00
14	1,2-Dichlorobenzene	1.479	1.494	-1.0	79	0.00
15	Benzyl Alcohol	1.162	1.091	6.1	72	-0.01
16	2,2'-oxybis(1-Chloropropane	1.879	1.787	4.9	73	-0.01
17	2-Methylphenol	1.163	1.142	1.8	75	-0.01
18	Hexachloroethane	0.560	0.561	-0.2	78	-0.01
19 P	n-Nitroso-di-n-propylamine	0.985	0.970	1.5	74	-0.01
20	3+4-Methylphenols	1.612	1.555	3.5	75	-0.01
21 I	Naphthalene-d8	1.000	1.000	0.0	78	-0.02
22	Acetophenone	0.508	0.495	2.6	76	-0.01
23 S	Nitrobenzene-d5	0.352	0.339	3.7	74	-0.01
24	Nitrobenzene	0.356	0.345	3.1	74	-0.01
25	Isophorone	0.698	0.678	2.9	74	-0.02
26 C	2-Nitrophenol	0.180	0.178	1.1	73	-0.02
27	2,4-Dimethylphenol	0.325	0.288	11.4	68	-0.01
28	bis(2-Chloroethoxy)methane	0.437	0.417	4.6	74	-0.02
29 C	2,4-Dichlorophenol	0.325	0.327	-0.6	77	-0.02
30	1,2,4-Trichlorobenzene	0.331	0.339	-2.4	80	-0.02
31	Naphthalene	1.081	1.077	0.4	78	-0.02
32	Benzoic acid	0.274	0.232	15.3	65	-0.02
33	4-Chloroaniline	0.460	0.444	3.5	73	-0.02
34 C	Hexachlorobutadiene	0.215	0.224	-4.2	81	-0.02
35	Caprolactam	0.116	0.113	2.6	74	-0.02
36 C	4-Chloro-3-methylphenol	0.345	0.343	0.6	75	-0.02
37	2-Methylnaphthalene	0.766	0.774	-1.0	78	-0.02
38	1-Methylnaphthalene	0.749	0.754	-0.7	77	-0.02
39 I	Acenaphthene-d10	1.000	1.000	0.0	78	-0.04
40	1,2,4,5-Tetrachlorobenzene	0.605	0.618	-2.1	81	-0.02
41 P	Hexachlorocyclopentadiene	0.396	0.331	16.4	65	-0.02
42 S	2,4,6-Tribromophenol	0.259	0.283	-9.3	84	-0.04
43 C	2,4,6-Trichlorophenol	0.415	0.419	-1.0	78	-0.02
44	2,4,5-Trichlorophenol	0.463	0.476	-2.8	80	-0.02
45 S	2-Fluorobiphenyl	1.365	1.389	-1.8	81	-0.03
46	1,1'-Biphenyl	1.506	1.486	1.3	78	-0.02
47	2-Chloronaphthalene	1.184	1.166	1.5	78	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112725\
 Data File : BP026190.D
 Acq On : 26 Nov 2025 14:04
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 14:52:55 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.329	0.309	6.1	72	-0.03
49	Acenaphthylene	1.953	1.945	0.4	78	-0.03
50	Dimethylphthalate	1.494	1.484	0.7	78	-0.03
51	2,6-Dinitrotoluene	0.296	0.306	-3.4	76	-0.03
52 C	Acenaphthene	1.145	1.138	0.6	78	-0.03
53	3-Nitroaniline	0.349	0.335	4.0	72	-0.03
54 P	2,4-Dinitrophenol	0.179	0.172	3.9	72	-0.04
55	Dibenzofuran	1.773	1.797	-1.4	80	-0.03
56 P	4-Nitrophenol	0.315	0.287	8.9	71	-0.04
57	2,4-Dinitrotoluene	0.442	0.457	-3.4	76	-0.04
58	Fluorene	1.401	1.433	-2.3	81	-0.04
59	2,3,4,6-Tetrachlorophenol	0.393	0.409	-4.1	80	-0.04
60	Diethylphthalate	1.505	1.489	1.1	77	-0.04
61	4-Chlorophenyl-phenylether	0.697	0.717	-2.9	81	-0.04
62	4-Nitroaniline	0.362	0.337	6.9	72	-0.03
63	Azobenzene	1.269	1.234	2.8	75	-0.04
64 I	Phenanthrene-d10	1.000	1.000	0.0	80	-0.04
65	4,6-Dinitro-2-methylphenol	0.127	0.125	1.6	75	-0.04
66 c	n-Nitrosodiphenylamine	0.619	0.613	1.0	79	-0.04
67	4-Bromophenyl-phenylether	0.221	0.229	-3.6	80	-0.04
68	Hexachlorobenzene	0.257	0.267	-3.9	81	-0.04
69	Atrazine	0.233	0.229	1.7	77	-0.04
70 C	Pentachlorophenol	0.180	0.185	-2.8	81	-0.04
71	Phenanthrene	1.127	1.121	0.5	80	-0.04
72	Anthracene	1.149	1.160	-1.0	79	-0.04
73	Carbazole	1.091	1.075	1.5	78	-0.03
74	Di-n-butylphthalate	1.294	1.301	-0.5	77	-0.04
75 C	Fluoranthene	1.369	1.384	-1.1	81	-0.03
76 I	Chrysene-d12	1.000	1.000	0.0	85	-0.02
77	Benzydine	0.635	0.574	9.6	79	-0.03
78	Pyrene	1.311	1.274	2.8	81	-0.02
79 S	Terphenyl-d14	0.981	0.995	-1.4	85	-0.03
80	Butylbenzylphthalate	0.578	0.555	4.0	78	-0.02
81	Benzo(a)anthracene	1.374	1.353	1.5	84	-0.02
82	3,3'-Dichlorobenzidine	0.500	0.471	5.8	80	-0.04
83	Chrysene	1.295	1.282	1.0	85	-0.02
84	Bis(2-ethylhexyl)phthalate	0.874	0.837	4.2	76	-0.02
85 c	Di-n-octyl phthalate	1.529	1.430	6.5	77	-0.04
86 I	Perylene-d12	1.000	1.000	0.0	85	-0.05
87	Indeno(1,2,3-cd)pyrene	1.500	1.559	-3.9	88	-0.10
88	Benzo(b)fluoranthene	1.218	1.260	-3.4	87	-0.05
89	Benzo(k)fluoranthene	1.217	1.237	-1.6	87	-0.05
90 C	Benzo(a)pyrene	1.154	1.171	-1.5	86	-0.05
91	Dibenzo(a,h)anthracene	1.228	1.272	-3.6	88	-0.11
92	Benzo(g,h,i)perylene	1.220	1.265	-3.7	88	-0.12

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112725\
Data File : BP026190.D
Acq On : 26 Nov 2025 14:04
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 26 14:52:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0