

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112725\
 Data File : BP026190.D
 Acq On : 26 Nov 2025 14:04
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 14:52:55 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	78	0.00
2	1,4-Dioxane	40.000	38.474	3.8	78	0.00
3	Pyridine	40.000	37.182	7.0	73	0.00
4	n-Nitrosodimethylamine	40.000	36.645	8.4	72	0.00
5 S	2-Fluorophenol	80.000	76.074	4.9	74	0.00
6	Aniline	40.000	37.655	5.9	73	0.00
7 S	Phenol-d6	80.000	77.469	3.2	74	-0.01
8	2-Chlorophenol	40.000	39.566	1.1	77	0.00
9	Benzaldehyde	40.000	41.356	-3.4	73	0.00
10 C	Phenol	40.000	38.377	4.1	73	0.00
11	bis(2-Chloroethyl)ether	40.000	38.745	3.1	74	0.00
12	1,3-Dichlorobenzene	40.000	40.666	-1.7	80	-0.01
13 C	1,4-Dichlorobenzene	40.000	40.343	-0.9	79	0.00
14	1,2-Dichlorobenzene	40.000	40.425	-1.1	79	0.00
15	Benzyl Alcohol	40.000	37.562	6.1	72	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	38.036	4.9	73	-0.01
17	2-Methylphenol	40.000	39.275	1.8	75	-0.01
18	Hexachloroethane	40.000	40.088	-0.2	78	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	39.388	1.5	74	-0.01
20	3+4-Methylphenols	40.000	38.595	3.5	75	-0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	78	-0.02
22	Acetophenone	40.000	38.955	2.6	76	-0.01
23 S	Nitrobenzene-d5	80.000	77.009	3.7	74	-0.01
24	Nitrobenzene	40.000	38.737	3.2	74	-0.01
25	Isophorone	40.000	38.843	2.9	74	-0.02
26 C	2-Nitrophenol	40.000	39.520	1.2	73	-0.02
27	2,4-Dimethylphenol	40.000	35.401	11.5	68	-0.01
28	bis(2-Chloroethoxy)methane	40.000	38.166	4.6	74	-0.02
29 C	2,4-Dichlorophenol	40.000	40.248	-0.6	77	-0.02
30	1,2,4-Trichlorobenzene	40.000	40.970	-2.4	80	-0.02
31	Naphthalene	40.000	39.846	0.4	78	-0.02
32	Benzoic acid	40.000	33.881	15.3	65	-0.02
33	4-Chloroaniline	40.000	38.647	3.4	73	-0.02
34 C	Hexachlorobutadiene	40.000	41.656	-4.1	81	-0.02
35	Caprolactam	40.000	38.718	3.2	74	-0.02
36 C	4-Chloro-3-methylphenol	40.000	39.714	0.7	75	-0.02
37	2-Methylnaphthalene	40.000	40.411	-1.0	78	-0.02
38	1-Methylnaphthalene	40.000	40.272	-0.7	77	-0.02
39 I	Acenaphthene-d10	20.000	20.000	0.0	78	-0.04
40	1,2,4,5-Tetrachlorobenzene	40.000	40.855	-2.1	81	-0.02
41 P	Hexachlorocyclopentadiene	40.000	33.460	16.3	65	-0.02
42 S	2,4,6-Tribromophenol	80.000	87.271	-9.1	84	-0.04
43 C	2,4,6-Trichlorophenol	40.000	40.449	-1.1	78	-0.02
44	2,4,5-Trichlorophenol	40.000	41.124	-2.8	80	-0.02
45 S	2-Fluorobiphenyl	80.000	81.419	-1.8	81	-0.03
46	1,1'-Biphenyl	40.000	39.470	1.3	78	-0.02
47	2-Chloronaphthalene	40.000	39.383	1.5	78	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112725\
 Data File : BP026190.D
 Acq On : 26 Nov 2025 14:04
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 14:52:55 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.522	6.2	72	-0.03
49	Acenaphthylene	40.000	39.825	0.4	78	-0.03
50	Dimethylphthalate	40.000	39.727	0.7	78	-0.03
51	2,6-Dinitrotoluene	40.000	41.338	-3.3	76	-0.03
52 C	Acenaphthene	40.000	39.764	0.6	78	-0.03
53	3-Nitroaniline	40.000	38.439	3.9	72	-0.03
54 P	2,4-Dinitrophenol	40.000	38.476	3.8	72	-0.04
55	Dibenzofuran	40.000	40.551	-1.4	80	-0.03
56 P	4-Nitrophenol	40.000	36.400	9.0	71	-0.04
57	2,4-Dinitrotoluene	40.000	41.321	-3.3	76	-0.04
58	Fluorene	40.000	40.906	-2.3	81	-0.04
59	2,3,4,6-Tetrachlorophenol	40.000	41.708	-4.3	80	-0.04
60	Diethylphthalate	40.000	39.588	1.0	77	-0.04
61	4-Chlorophenyl-phenylether	40.000	41.145	-2.9	81	-0.04
62	4-Nitroaniline	40.000	37.250	6.9	72	-0.03
63	Azobenzene	40.000	38.915	2.7	75	-0.04
64 I	Phenanthrene-d10	20.000	20.000	0.0	80	-0.04
65	4,6-Dinitro-2-methylphenol	40.000	39.438	1.4	75	-0.04
66 c	n-Nitrosodiphenylamine	40.000	39.653	0.9	79	-0.04
67	4-Bromophenyl-phenylether	40.000	41.382	-3.5	80	-0.04
68	Hexachlorobenzene	40.000	41.599	-4.0	81	-0.04
69	Atrazine	40.000	39.304	1.7	77	-0.04
70 C	Pentachlorophenol	40.000	41.163	-2.9	81	-0.04
71	Phenanthrene	40.000	39.799	0.5	80	-0.04
72	Anthracene	40.000	40.378	-0.9	79	-0.04
73	Carbazole	40.000	39.386	1.5	78	-0.03
74	Di-n-butylphthalate	40.000	40.204	-0.5	77	-0.04
75 C	Fluoranthene	40.000	40.445	-1.1	81	-0.03
76 I	Chrysene-d12	20.000	20.000	0.0	85	-0.02
77	Benzydine	40.000	36.195	9.5	79	-0.03
78	Pyrene	40.000	38.873	2.8	81	-0.02
79 S	Terphenyl-d14	80.000	81.136	-1.4	85	-0.03
80	Butylbenzylphthalate	40.000	38.429	3.9	78	-0.02
81	Benzo(a)anthracene	40.000	39.415	1.5	84	-0.02
82	3,3'-Dichlorobenzidine	40.000	37.690	5.8	80	-0.04
83	Chrysene	40.000	39.612	1.0	85	-0.02
84	Bis(2-ethylhexyl)phthalate	40.000	38.312	4.2	76	-0.02
85 c	Di-n-octyl phthalate	40.000	37.427	6.4	77	-0.04
86 I	Perylene-d12	20.000	20.000	0.0	85	-0.05
87	Indeno(1,2,3-cd)pyrene	40.000	41.575	-3.9	88	-0.10
88	Benzo(b)fluoranthene	40.000	41.380	-3.5	87	-0.05
89	Benzo(k)fluoranthene	40.000	40.681	-1.7	87	-0.05
90 C	Benzo(a)pyrene	40.000	40.585	-1.5	86	-0.05
91	Dibenzo(a,h)anthracene	40.000	41.444	-3.6	88	-0.11
92	Benzo(g,h,i)perylene	40.000	41.449	-3.6	88	-0.12

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112725\
Data File : BP026190.D
Acq On : 26 Nov 2025 14:04
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 26 14:52:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0