

Surrogate Summary

SW-846

SDG No.: Q3483

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB170598BL	PB170598BL	2-Fluorophenol	150	136	91		10	105
		Phenol-d6	150	136	90		10	103
		Nitrobenzene-d5	100	85.1	85		10	108
		2-Fluorobiphenyl	100	85.3	85		10	108
		2,4,6-Tribromophenol	150	129	86		10	116
		Terphenyl-d14	100	89.0	89		10	109
PB170598BS	PB170598BS	2-Fluorophenol	150	130	86		10	105
		Phenol-d6	150	131	87		10	103
		Nitrobenzene-d5	100	78.2	78		10	108
		2-Fluorobiphenyl	100	75.9	76		10	108
		2,4,6-Tribromophenol	150	126	84		10	116
		Terphenyl-d14	100	83.7	84		10	109
Q3483-11	HW1025-PT-BNA-SOIL	2-Fluorophenol	150	82.0	55		10	105
		Phenol-d6	150	86.4	58		10	103
		Nitrobenzene-d5	100	57.2	57		10	108
		2-Fluorobiphenyl	100	68.2	68		10	108
		2,4,6-Tribromophenol	150	128	85		10	116
		Terphenyl-d14	100	87.7	88		10	109
Q3483-11DL	HW1025-PT-BNA-SOILDL	2-Fluorophenol	150	69.6	46		10	105
		Phenol-d6	150	73.3	49		10	103
		Nitrobenzene-d5	100	46.7	47		10	108
		2-Fluorobiphenyl	100	64.5	65		10	108
		2,4,6-Tribromophenol	150	113	76		10	116
		Terphenyl-d14	100	81.7	82		10	109
Q3658-01MS	NB-07-111725MS	2-Fluorophenol	150	98.6	66		10	105
		Phenol-d6	150	99.7	66		10	103
		Nitrobenzene-d5	100	62.3	62		10	108
		2-Fluorobiphenyl	100	59.3	59		10	108
		2,4,6-Tribromophenol	150	90.8	61		10	116
		Terphenyl-d14	100	73.4	73		10	109
Q3658-01MSD	NB-07-111725MSD	2-Fluorophenol	150	98.7	66		10	105
		Phenol-d6	150	99.1	66		10	103
		Nitrobenzene-d5	100	61.9	62		10	108
		2-Fluorobiphenyl	100	57.3	57		10	108
		2,4,6-Tribromophenol	150	93.0	62		10	116
		Terphenyl-d14	100	72.9	73		10	109