

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3483

Client ID : SSTDCCC040

Date Analyzed: 11/26/2025

Lab File ID: BP026190.D

Time Analyzed: 14:04

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	263413	7.666	1101920	10.43	729934	14.29
UPPER LIMIT	526826	8.166	2203840	10.925	1459870	14.789
LOWER LIMIT	131707	7.166	550960	9.925	364967	13.789
EPA SAMPLE NO.						
01 HW1025-PT-BNA-SOIL	262012	7.66	1065580	10.43	680012	14.30
02 HW1025-PT-BNA-SOILDL	290064	7.66	1229870	10.42	837371	14.29

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.