

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BG064744.D
Instrument ID: BNA_G

Contract: ALLI03
SDG NO.: Q3483
DFTPP Injection Date: 11/12/2025
DFTPP Injection Time: 10:37

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.2 (0.6) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.1 (0.2) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.7
365	Greater than 1% of mass 198	5.6
441	Present, but less than mass 443	84.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BG064745.D	11/12/2025	11:17
SSTDICC005	SSTDICC005	BG064746.D	11/12/2025	11:58
SSTDICC010	SSTDICC010	BG064747.D	11/12/2025	12:39
SSTDICC020	SSTDICC020	BG064748.D	11/12/2025	13:20
SSTDICCC040	SSTDICCC040	BG064749.D	11/12/2025	14:00
SSTDICC050	SSTDICC050	BG064750.D	11/12/2025	14:41
SSTDICC060	SSTDICC060	BG064751.D	11/12/2025	15:23
SSTDICC080	SSTDICC080	BG064752.D	11/12/2025	16:04