

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP026142.D
Instrument ID: BNA_P

Contract: ALLI03
SDG NO.: Q3483
DFTPP Injection Date: 11/18/2025
DFTPP Injection Time: 13:45

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.6) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.1 (0.4) 1
197	Less than 2.0% of mass 198	0.1
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	79.0
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.3 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP026143.D	11/18/2025	14:31
SSTDICC005	SSTDICC005	BP026144.D	11/18/2025	15:12
SSTDICC010	SSTDICC010	BP026145.D	11/18/2025	15:54
SSTDICC020	SSTDICC020	BP026146.D	11/18/2025	16:35
SSTDICCC040	SSTDICCC040	BP026147.D	11/18/2025	17:57
SSTDICC050	SSTDICC050	BP026148.D	11/18/2025	18:38
SSTDICC060	SSTDICC060	BP026149.D	11/18/2025	20:00
SSTDICC080	SSTDICC080	BP026150.D	11/18/2025	21:22